

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011023\
 Data File : VY012147.D
 Acq On : 10 Jan 2023 17:23
 Operator : KP/MD
 Sample : 01106-04
 Misc : 5.90g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-B

Quant Time: Jan 11 03:10:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 07 01:46:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	74992	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	126861	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	124550	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	55996	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	50433	59.850	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.700%
35) Dibromofluoromethane	7.722	113	41679	49.985	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.960%
50) Toluene-d8	10.179	98	153722	46.707	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	93.420%
62) 4-Bromofluorobenzene	12.483	95	46589	40.498	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	81.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011023\
Data File : VY012147.D
Acq On : 10 Jan 2023 17:23
Operator : KP/MD
Sample : 01106-04
Misc : 5.90g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-B

Quant Time: Jan 11 03:10:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
Quant Title : SW846 8260
QLast Update : Sat Jan 07 01:46:48 2023
Response via : Initial Calibration

