

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011123\
 Data File : VY012154.D
 Acq On : 11 Jan 2023 11:41
 Operator : KP/MD
 Sample : 01089-01RE
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PL-1-010923RE

Quant Time: Jan 11 13:04:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 07 01:46:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	5637	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	8817	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	8156	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	2793	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	3904	61.635	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	123.260%
35) Dibromofluoromethane	7.721	113	1661	28.661	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	57.320%
50) Toluene-d8	10.184	98	10854	47.451	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.900%
62) 4-Bromofluorobenzene	12.489	95	2899	36.258	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	72.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011123\
Data File : VY012154.D
Acq On : 11 Jan 2023 11:41
Operator : KP/MD
Sample : 01089-01RE
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PL-1-010923RE

Quant Time: Jan 11 13:04:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
Quant Title : SW846 8260
QLast Update : Sat Jan 07 01:46:48 2023
Response via : Initial Calibration

