

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011123\
 Data File : VY012159.D
 Acq On : 11 Jan 2023 13:38
 Operator : KP/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Jan 12 05:49:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 07 01:46:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	231665	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	396771	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	377404	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	181191	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	127176	48.855	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.720%
35) Dibromofluoromethane	7.716	113	117424	45.026	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.060%
50) Toluene-d8	10.179	98	473767	46.026	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.060%
62) 4-Bromofluorobenzene	12.483	95	163934	45.562	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011123\
Data File : VY012159.D
Acq On : 11 Jan 2023 13:38
Operator : KP/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Jan 12 05:49:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
Quant Title : SW846 8260
QLast Update : Sat Jan 07 01:46:48 2023
Response via : Initial Calibration

