

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011123\
 Data File : VY012161.D
 Acq On : 11 Jan 2023 14:25
 Operator : KP/MD
 Sample : 01106-01
 Misc : 5.95g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-3

Quant Time: Jan 12 05:49:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 07 01:46:48 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	212491	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	361261	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	344266	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	161664	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	120746	50.570	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	101.140%	
35) Dibromofluoromethane	7.716	113	109110	45.951	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	91.900%	
50) Toluene-d8	10.179	98	427238	45.585	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	91.180%	
62) 4-Bromofluorobenzene	12.483	95	144303	44.048	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	88.100%	
Target Compounds						
20) Methylene Chloride	4.716	84	29903	7.572	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011123\
Data File : VY012161.D
Acq On : 11 Jan 2023 14:25
Operator : KP/MD
Sample : 01106-01
Misc : 5.95g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH-3

Quant Time: Jan 12 05:49:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
Quant Title : SW846 8260
QLast Update : Sat Jan 07 01:46:48 2023
Response via : Initial Calibration

