

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011123\
 Data File : VY012162.D
 Acq On : 11 Jan 2023 14:48
 Operator : KP/MD
 Sample : 01106-04RE
 Misc : 5.94g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-B

Quant Time: Jan 12 05:50:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 07 01:46:48 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	195160	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	330494	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	317962	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	150372	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	115235	52.548	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	105.100%	
35) Dibromofluoromethane	7.715	113	101865	46.893	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	93.780%	
50) Toluene-d8	10.178	98	393713	45.919	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	91.840%	
62) 4-Bromofluorobenzene	12.483	95	134481	44.872	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	89.740%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.716	84	25712	6.661	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011123\
Data File : VY012162.D
Acq On : 11 Jan 2023 14:48
Operator : KP/MD
Sample : 01106-04RE
Misc : 5.94g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-B

Quant Time: Jan 12 05:50:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
Quant Title : SW846 8260
QLast Update : Sat Jan 07 01:46:48 2023
Response via : Initial Calibration

