



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011123\
Data File : VY012149.D
Acq On : 11 Jan 2023 08:53
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
Title : SW846 8260
Last Update : Sat Jan 07 01:46:48 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	16.4	10679	PASS
75	95	30	60	52.0	33824	PASS
95	95	100	100	100.0	65080	PASS
96	95	5	9	6.8	4396	PASS
173	174	0.00	2	1.0	613	PASS
174	95	50	100	92.2	59984	PASS
175	174	5	9	7.5	4508	PASS
176	174	95	101	98.6	59131	PASS
177	176	5	9	6.5	3821	PASS