

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011221\
 Data File : VY003899.D
 Acq On : 12 Jan 2021 13:34
 Operator : SY/MD
 Sample : VY0112SBL01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0112SBL01

Quant Time: Jan 13 02:32:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 12 11:47:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	146158	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	215991	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	200451	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	103824	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	116398	51.27	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.54%
35) Dibromofluoromethane	7.721	113	90088	51.04	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.08%
50) Toluene-d8	10.178	98	365681	51.12	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.24%
62) 4-Bromofluorobenzene	12.477	95	126924	48.58	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.16%
Target Compounds						
20) Methylene Chloride	4.716	84	16912	8.04	ug/l	Qvalue # 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011221\
Data File : VY003899.D
Acq On : 12 Jan 2021 13:34
Operator : SY/MD
Sample : VY0112SBL01
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0112SBL01

Quant Time: Jan 13 02:32:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011221S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 12 11:47:10 2021
Response via : Initial Calibration

