

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011221\
 Data File : VY003909.D
 Acq On : 12 Jan 2021 18:26
 Operator : SY/MD
 Sample : M1075-01
 Misc : 5.58G/5ML/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 72-11977

Quant Time: Jan 13 02:35:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 12 11:47:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	150302	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	253299	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	248998	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	130964	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	101133	43.32	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	86.64%
35) Dibromofluoromethane	7.722	113	81152	39.20	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	78.40%
50) Toluene-d8	10.179	98	326609	38.93	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	77.86%
62) 4-Bromofluorobenzene	12.477	95	120678	39.39	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	78.78%
Target Compounds						
95) Naphthalene	15.233	128	174878	26.00	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011221\
Data File : VY003909.D
Acq On : 12 Jan 2021 18:26
Operator : SY/MD
Sample : M1075-01
Misc : 5.58G/5ML/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
72-11977

Quant Time: Jan 13 02:35:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011221S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 12 11:47:10 2021
Response via : Initial Calibration

