

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011322\
 Data File : VY007143.D
 Acq On : 13 Jan 2022 15:09
 Operator : SY/MD
 Sample : VY0113SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0113SBL01

Quant Time: Jan 14 05:52:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	118043	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	216427	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	204663	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	87442	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	74657	51.544	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.080%
35) Dibromofluoromethane	7.716	113	71302	48.722	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.440%
50) Toluene-d8	10.179	98	266544	49.846	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.700%
62) 4-Bromofluorobenzene	12.477	95	96438	50.700	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011322\
Data File : VY007143.D
Acq On : 13 Jan 2022 15:09
Operator : SY/MD
Sample : VY0113SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0113SBL01

Quant Time: Jan 14 05:52:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 14 05:48:29 2022
Response via : Initial Calibration

