

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011322\
 Data File : VY007146.D
 Acq On : 13 Jan 2022 16:27
 Operator : SY/MD
 Sample : N1099-09
 Misc : 5.72/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VACLOT-BACKFILL-TILCON-09

Quant Time: Jan 14 05:53:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	110582	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	207831	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	199584	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	87658	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	75503	55.646	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.300%
35) Dibromofluoromethane	7.716	113	69094	49.166	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.340%
50) Toluene-d8	10.179	98	254155	49.495	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.980%
62) 4-Bromofluorobenzene	12.483	95	96595	52.883	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.760%
Target Compounds						
16) Acetone	3.942	43	17387	49.376	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011322\
Data File : VY007146.D
Acq On : 13 Jan 2022 16:27
Operator : SY/MD
Sample : N1099-09
Misc : 5.72/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VACLOT-BACKFILL-TILCON-09

Quant Time: Jan 14 05:53:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 14 05:48:29 2022
Response via : Initial Calibration

