

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011322\
 Data File : VY007148.D
 Acq On : 13 Jan 2022 18:19
 Operator : SY/MD
 Sample : N1119-02
 Misc : 6.41g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NORTH-EAST-CORNER-EXCAVATION-SOUTH-SIDEWALL

Quant Time: Jan 14 05:54:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	112500	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	200099	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	185274	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	82854	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	80437	58.271	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.540%
35) Dibromofluoromethane	7.715	113	66844	49.403	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.800%
50) Toluene-d8	10.178	98	242054	48.960	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.920%
62) 4-Bromofluorobenzene	12.483	95	90423	51.417	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	102.840%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	4.960	59	1684	9.681	ug/l #	73
16) Acetone	3.942	43	12363	34.510	ug/l #	68
52) Toluene	10.245	92	4494	1.230	ug/l	99
68) m/p-Xylenes	11.696	106	4748	1.678	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011322\
Data File : VY007148.D
Acq On : 13 Jan 2022 18:19
Operator : SY/MD
Sample : N1119-02
Misc : 6.41g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NORTH-EAST-CORNER-EXCAVATION-SOUTH-SIDEWALL

Quant Time: Jan 14 05:54:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 14 05:48:29 2022
Response via : Initial Calibration

