

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011322\
 Data File : VY007154.D
 Acq On : 13 Jan 2022 21:00
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 6.79g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 14 06:11:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.175	0.300	-71.4#	162#	0.00
3 P	Chloromethane	0.368	0.364	1.1	103	0.00
4 C	Vinyl Chloride	0.535	0.535	0.0	95	0.00
5 T	Bromomethane	0.346	0.369	-6.6	96	0.00
6 T	Chloroethane	0.334	0.344	-3.0	97	0.00
7 T	Trichlorofluoromethane	0.577	0.571	1.0	95	0.00
8 T	Diethyl Ether	0.318	0.335	-5.3	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.562	0.564	-0.4	96	0.00
10 T	Methyl Iodide	0.831	0.831	0.0	95	0.00
11 T	Tert butyl alcohol	0.058	0.055	5.2	104	0.00
12 CM	1,1-Dichloroethene	0.566	0.564	0.4#	97	0.00
13 T	Acrolein	0.015	0.011	26.7#	95	0.00
14 T	Allyl chloride	0.907	0.925	-2.0	97	0.00
15 T	Acrylonitrile	0.153	0.166	-8.5	101	-0.01
16 T	Acetone	0.159	0.156	1.9	93	0.00
17 T	Carbon Disulfide	1.571	1.541	1.9	95	0.00
18 T	Methyl Acetate	0.567	0.568	-0.2	102	0.00
19 T	Methyl tert-butyl Ether	1.114	1.139	-2.2	97	0.00
20 T	Methylene Chloride	0.658	0.626	4.9	98	0.00
21 T	trans-1,2-Dichloroethene	0.613	0.624	-1.8	98	0.00
22 T	Diisopropyl ether	1.838	1.901	-3.4	97	0.00
23 T	Vinyl Acetate	1.081	1.179	-9.1	99	0.00
24 P	1,1-Dichloroethane	1.118	1.159	-3.7	98	-0.01
25 T	2-Butanone	0.206	0.220	-6.8	101	0.00
26 T	2,2-Dichloropropane	0.815	0.763	6.4	91	0.00
27 T	cis-1,2-Dichloroethene	0.696	0.724	-4.0	99	0.00
28 T	Bromochloromethane	0.404	0.459	-13.6	98	0.00
29 T	Tetrahydrofuran	0.132	0.145	-9.8	104	0.00
30 C	Chloroform	1.162	1.222	-5.2#	100	0.00
31 T	Cyclohexane	1.037	0.996	4.0	97	0.00
32 T	1,1,1-Trichloroethane	0.993	1.016	-2.3	97	0.00
33 S	1,2-Dichloroethane-d4	0.614	0.636	-3.6	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.338	0.345	-2.1	100	0.00
36 T	1,1-Dichloropropene	0.540	0.539	0.2	98	0.00
37 T	Ethyl Acetate	0.277	0.294	-6.1	100	0.00
38 T	Carbon Tetrachloride	0.568	0.570	-0.4	98	0.00
39 T	Methylcyclohexane	0.659	0.661	-0.3	98	0.00
40 TM	Benzene	1.440	1.477	-2.6	99	0.00
41 T	Methacrylonitrile	0.160	0.196	-22.5	131	0.01
42 TM	1,2-Dichloroethane	0.454	0.481	-5.9	101	0.00
43 T	Isopropyl Acetate	0.539	0.585	-8.5	103	0.00
44 TM	Trichloroethene	0.390	0.390	0.0	96	0.00
45 C	1,2-Dichloropropane	0.366	0.379	-3.6#	100	0.00
46 T	Dibromomethane	0.217	0.228	-5.1	101	0.00
47 T	Bromodichloromethane	0.532	0.560	-5.3	100	0.00
48 T	Methyl methacrylate	0.245	0.261	-6.5	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011322\
 Data File : VY007154.D
 Acq On : 13 Jan 2022 21:00
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 6.79g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 14 06:11:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.003	-50.0#	107	0.00
50 S	Toluene-d8	1.235	1.266	-2.5	101	0.00
51 T	4-Methyl-2-Pentanone	0.273	0.297	-8.8	103	0.00
52 CM	Toluene	0.913	0.929	-1.8#	96	0.00
53 T	t-1,3-Dichloropropene	0.562	0.593	-5.5	99	0.00
54 T	cis-1,3-Dichloropropene	0.629	0.650	-3.3	97	0.00
55 T	1,1,2-Trichloroethane	0.294	0.308	-4.8	100	0.00
56 T	Ethyl methacrylate	0.414	0.448	-8.2	102	0.00
57 T	1,3-Dichloropropane	0.512	0.546	-6.6	102	0.00
58 T	2-Chloroethyl Vinyl ether	0.188	0.186	1.1	98	0.00
59 T	2-Hexanone	0.186	0.204	-9.7	106	0.00
60 T	Dibromochloromethane	0.361	0.382	-5.8	100	0.00
61 T	1,2-Dibromoethane	0.289	0.310	-7.3	101	0.00
62 S	4-Bromofluorobenzene	0.439	0.456	-3.9	103	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.360	0.370	-2.8	102	0.00
65 PM	Chlorobenzene	1.093	1.123	-2.7	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.405	0.418	-3.2	98	0.00
67 C	Ethyl Benzene	2.055	2.101	-2.2#	99	0.00
68 T	m/p-Xylenes	0.764	0.778	-1.8	99	0.00
69 T	o-Xylene	0.719	0.748	-4.0	101	0.00
70 T	Styrene	1.213	1.278	-5.4	100	0.00
71 P	Bromoform	0.234	0.247	-5.6	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	4.319	4.322	-0.1	99	0.00
74 T	N-amyl acetate	1.155	1.278	-10.6	108	0.00
75 P	1,1,2,2-Tetrachloroethane	0.840	0.881	-4.9	103	0.00
76 T	1,2,3-Trichloropropane	0.570	0.656	-15.1	100	0.00
77 T	Bromobenzene	0.940	0.957	-1.8	101	0.00
78 T	n-propylbenzene	5.353	5.356	-0.1	100	0.00
79 T	2-Chlorotoluene	2.977	3.030	-1.8	101	0.00
80 T	1,3,5-Trimethylbenzene	3.546	3.586	-1.1	101	0.00
81 T	trans-1,4-Dichloro-2-butene	0.311	0.324	-4.2	99	0.00
82 T	4-Chlorotoluene	3.105	3.120	-0.5	99	0.00
83 T	tert-Butylbenzene	3.064	3.108	-1.4	100	0.00
84 T	1,2,4-Trimethylbenzene	3.515	3.585	-2.0	101	0.00
85 T	sec-Butylbenzene	4.584	4.553	0.7	99	0.00
86 T	p-Isopropyltoluene	3.767	3.801	-0.9	100	0.00
87 T	1,3-Dichlorobenzene	1.839	1.881	-2.3	102	0.00
88 T	1,4-Dichlorobenzene	1.849	1.859	-0.5	101	0.00
89 T	n-Butylbenzene	3.709	3.718	-0.2	100	0.00
90 T	Hexachloroethane	0.736	0.756	-2.7	101	0.00
91 T	1,2-Dichlorobenzene	1.637	1.674	-2.3	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.145	0.151	-4.1	106	0.00
93 T	1,2,4-Trichlorobenzene	0.995	1.001	-0.6	101	0.00
94 T	Hexachlorobutadiene	0.513	0.507	1.2	101	0.00
95 T	Naphthalene	2.114	2.263	-7.0	108	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011322\
Data File : VY007154.D
Acq On : 13 Jan 2022 21:00
Operator : SY/MD
Sample : VSTDCCC050
Misc : 6.79g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jan 14 06:11:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 14 05:48:29 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.849	0.874	-2.9	104	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 5