

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011425\
 Data File : VY020815.D
 Acq On : 14 Jan 2025 11:47
 Operator : SY/MD
 Sample : Q1074-01
 Misc : 4.44g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HACKENSACK-DGA

Quant Time: Jan 15 03:27:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 01:31:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	174413	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.628	114	281645	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.426	117	220855	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.359	152	79917	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	96200	64.893	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	129.780%
35) Dibromofluoromethane	7.646	113	89546	50.749	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.500%
50) Toluene-d8	10.115	98	314789	51.177	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.360%
62) 4-Bromofluorobenzene	12.414	95	89562	39.429	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	78.860%
Target Compounds						
51) 4-Methyl-2-Pentanone	10.006	43	1065	1.170	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011425\
Data File : VY020815.D
Acq On : 14 Jan 2025 11:47
Operator : SY/MD
Sample : Q1074-01
Misc : 4.44g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HACKENSACK-DGA

Quant Time: Jan 15 03:27:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 01:31:46 2025
Response via : Initial Calibration

