

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011425\
 Data File : VY020823.D
 Acq On : 14 Jan 2025 15:38
 Operator : SY/MD
 Sample : Q1077-01
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-05-01132025

Quant Time: Jan 15 03:30:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 01:31:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.720	168	40	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.634	114	237	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.432	117	176	50.000	ug/l	# 0.01
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.35
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	41	120.593	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 241.180%#	
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	= 0.000%#	
50) Toluene-d8	10.115	98	161	31.105	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 62.220%	
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	29 - 146	Recovery	= 0.000%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011425\
Data File : VY020823.D
Acq On : 14 Jan 2025 15:38
Operator : SY/MD
Sample : Q1077-01
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-05-01132025

Quant Time: Jan 15 03:30:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 01:31:46 2025
Response via : Initial Calibration

