



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011425\
Data File : VY020809.D
Acq On : 14 Jan 2025 08:17
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
Title : SW846 8260
Last Update : Tue Jan 07 01:31:46 2025



AutoFind: Scans 1754, 1755, 1756; Background Corrected with Scan 1746

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.4	18486	PASS
75	95	30	60	50.0	56421	PASS
95	95	100	100	100.0	112765	PASS
96	95	5	9	6.6	7398	PASS
173	174	0.00	2	0.6	533	PASS
174	95	50	100	84.2	94955	PASS
175	174	5	9	7.7	7286	PASS
176	174	95	101	99.1	94080	PASS
177	176	5	9	6.9	6463	PASS