

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011623\
 Data File : VY012206.D
 Acq On : 16 Jan 2023 17:55
 Operator : KP/MD
 Sample : 01159-01
 Misc : 4.91g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-E

Quant Time: Jan 17 04:50:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	189153	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	317654	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	299902	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	139520	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	109553	71.570	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	143.140%
35) Dibromofluoromethane	7.716	113	93929	55.844	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.680%
50) Toluene-d8	10.179	98	373025	65.337	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	130.680%
62) 4-Bromofluorobenzene	12.483	95	123917	52.262	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011623\
Data File : VY012206.D
Acq On : 16 Jan 2023 17:55
Operator : KP/MD
Sample : 01159-01
Misc : 4.91g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-E

Quant Time: Jan 17 04:50:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

