

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011722\
 Data File : VY007179.D
 Acq On : 17 Jan 2022 10:25
 Operator : SY/MD
 Sample : VY0117SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0117SBL01

Quant Time: Jan 17 17:45:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	124316	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	226493	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	208698	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	90136	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	74111	48.585	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.180%
35) Dibromofluoromethane	7.721	113	74105	48.387	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.780%
50) Toluene-d8	10.184	98	276699	49.445	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.900%
62) 4-Bromofluorobenzene	12.483	95	97874	49.168	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011722\
Data File : VY007179.D
Acq On : 17 Jan 2022 10:25
Operator : SY/MD
Sample : VY0117SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0117SBL01

Quant Time: Jan 17 17:45:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 14 05:48:29 2022
Response via : Initial Calibration

