

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011722\
 Data File : VY007196.D
 Acq On : 17 Jan 2022 19:11
 Operator : SY/MD
 Sample : N1166-01
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SU-01-011422

Quant Time: Jan 18 03:11:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	1777	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	4318	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	5042	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	2179	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	2474	113.465	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	226.940%#
35) Dibromofluoromethane	7.722	113	1755	60.107	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	120.220%
50) Toluene-d8	10.185	98	5202	48.760	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.520%
62) 4-Bromofluorobenzene	12.477	95	2749	72.437	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	144.880%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011722\
Data File : VY007196.D
Acq On : 17 Jan 2022 19:11
Operator : SY/MD
Sample : N1166-01
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SU-01-011422

Quant Time: Jan 18 03:11:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 14 05:48:29 2022
Response via : Initial Calibration

