

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011722\
 Data File : VY007198.D
 Acq On : 17 Jan 2022 20:09
 Operator : SY/MD
 Sample : N1144-02RE
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RO-ETGI-352RE

Quant Time: Jan 18 03:11:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	31098	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	48740	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	33939	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	12902	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	17441	45.708	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.420%
35) Dibromofluoromethane	7.722	113	15604	47.346	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.700%
50) Toluene-d8	10.179	98	55131	45.781	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	91.560%
62) 4-Bromofluorobenzene	12.483	95	14937	34.870	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	69.740%
Target Compounds						
16) Acetone	3.942	43	1273	12.855	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011722\
Data File : VY007198.D
Acq On : 17 Jan 2022 20:09
Operator : SY/MD
Sample : N1144-02RE
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RO-ETGI-352RE

Quant Time: Jan 18 03:11:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 14 05:48:29 2022
Response via : Initial Calibration

