

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011822\
 Data File : VY007206.D
 Acq On : 18 Jan 2022 13:12
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Jan 19 00:31:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	103988	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	196236	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	183047	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	79959	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	68354	53.571	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 107.140%	
35) Dibromofluoromethane	7.722	113	64130	48.330	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 96.660%	
50) Toluene-d8	10.185	98	243369	50.195	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 100.380%	
62) 4-Bromofluorobenzene	12.483	95	87416	50.685	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 101.380%	
Target Compounds						
30) Chloroform	7.509	83	5073	2.100	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011822\
Data File : VY007206.D
Acq On : 18 Jan 2022 13:12
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Jan 19 00:31:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 14 05:48:29 2022
Response via : Initial Calibration

