

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012225\
 Data File : VY020894.D
 Acq On : 22 Jan 2025 11:27
 Operator : SY/MD
 Sample : Q1141-02
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SU-4-012125-VOC

Quant Time: Jan 23 00:44:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 23 00:38:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.731	168	440191	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.640	114	670720	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.438	117	523054	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.365	152	214524	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.085	65	164326	43.920	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	87.840%
35) Dibromofluoromethane	7.664	113	184197	43.836	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	87.680%
50) Toluene-d8	10.127	98	698835	47.708	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.420%
62) 4-Bromofluorobenzene	12.426	95	201670	37.282	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	74.560%
Target Compounds						
95) Naphthalene	15.169	128	74048	11.728	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012225\
Data File : VY020894.D
Acq On : 22 Jan 2025 11:27
Operator : SY/MD
Sample : Q1141-02
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SU-4-012125-VOC

Quant Time: Jan 23 00:44:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 23 00:38:42 2025
Response via : Initial Calibration

