



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012425\
Data File : VY020922.D
Acq On : 24 Jan 2025 08:44
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
Title : SW846 8260
Last Update : Thu Jan 23 00:38:42 2025



AutoFind: Scans 1755, 1756, 1757; Background Corrected with Scan 1747

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.3	16398	PASS
75	95	30	60	50.3	50720	PASS
95	95	100	100	100.0	100883	PASS
96	95	5	9	6.3	6375	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	90.9	91696	PASS
175	174	5	9	7.7	7103	PASS
176	174	95	101	96.3	88291	PASS
177	176	5	9	7.5	6630	PASS