

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012522\
 Data File : VY007294.D
 Acq On : 25 Jan 2022 16:52
 Operator : SY/MD
 Sample : N1266-04
 Misc : 6.93g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RB-10-(5-6)

Quant Time: Jan 26 09:54:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	82661	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	152780	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	144032	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	62469	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	52574	51.835	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.660%
35) Dibromofluoromethane	7.716	113	51042	49.408	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.820%
50) Toluene-d8	10.179	98	186802	49.487	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.980%
62) 4-Bromofluorobenzene	12.483	95	67124	49.990	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012522\
Data File : VY007294.D
Acq On : 25 Jan 2022 16:52
Operator : SY/MD
Sample : N1266-04
Misc : 6.93g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RB-10-(5-6)

Quant Time: Jan 26 09:54:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 14 05:48:29 2022
Response via : Initial Calibration

