

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012524\
 Data File : VY017101.D
 Acq On : 25 Jan 2024 11:27
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICSVY012524

Quant Time: Jan 26 04:40:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012424S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 25 04:44:10 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	0.434	0.331	23.7	89	0.00
3 P	Chloromethane	0.729	0.628	13.9	95	0.00
4 C	Vinyl Chloride	0.833	0.764	8.3#	98	0.00
5 T	Bromomethane	0.621	0.557	10.3	99	0.00
6 T	Chloroethane	0.534	0.505	5.4	99	0.00
7 T	Trichlorofluoromethane	0.966	0.796	17.6	92	0.00
8 T	Diethyl Ether	0.288	0.261	9.4	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.559	0.482	13.8	88	0.00
10 T	Methyl Iodide	0.531	0.505	4.9	88	0.00
11 T	Tert butyl alcohol	0.037	0.034	8.1	99	0.00
12 CM	1,1-Dichloroethene	0.512	0.460	10.2#	91	0.00
13 T	Acrolein	0.058	0.056	3.4	94	0.00
14 T	Allyl chloride	0.774	0.729	5.8	92	0.00
15 T	Acrylonitrile	0.125	0.118	5.6	92	-0.01
16 T	Acetone	0.158	0.119	24.7	94	0.00
17 T	Carbon Disulfide	1.543	1.364	11.6	89	0.00
18 T	Methyl Acetate	0.500	0.505	-1.0	95	-0.01
19 T	Methyl tert-butyl Ether	1.249	1.188	4.9	93	0.00
20 T	Methylene Chloride	0.671	0.531	20.9	90	0.00
21 T	trans-1,2-Dichloroethene	0.592	0.543	8.3	90	0.00
22 T	Diisopropyl ether	1.708	1.639	4.0	92	0.00
23 T	Vinyl Acetate	0.829	0.783	5.5	92	0.00
24 P	1,1-Dichloroethane	1.042	0.943	9.5	91	0.00
25 T	2-Butanone	0.194	0.160	17.5	93	0.00
26 T	2,2-Dichloropropane	0.890	0.818	8.1	92	0.00
27 T	cis-1,2-Dichloroethene	0.656	0.609	7.2	91	0.00
28 T	Bromochloromethane	0.415	0.449	-8.2	91	0.00
29 T	Tetrahydrofuran	0.099	0.096	3.0	92	0.00
30 C	Chloroform	1.061	0.966	9.0#	91	0.00
31 T	Cyclohexane	0.963	0.830	13.8	89	0.00
32 T	1,1,1-Trichloroethane	0.910	0.840	7.7	92	0.00
33 S	1,2-Dichloroethane-d4	0.487	0.512	-5.1	97	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.302	0.311	-3.0	94	0.00
36 T	1,1-Dichloropropene	0.497	0.460	7.4	92	0.00
37 T	Ethyl Acetate	0.213	0.201	5.6	96	0.00
38 T	Carbon Tetrachloride	0.487	0.440	9.7	91	0.00
39 T	Methylcyclohexane	0.597	0.542	9.2	87	0.00
40 TM	Benzene	1.442	1.328	7.9	90	0.00
41 T	Methacrylonitrile	0.114	0.103	9.6	98	0.00
42 TM	1,2-Dichloroethane	0.374	0.344	8.0	91	0.00
43 T	Isopropyl Acetate	0.399	0.377	5.5	92	0.00
44 TM	Trichloroethene	0.370	0.337	8.9	91	0.00
45 C	1,2-Dichloropropane	0.354	0.330	6.8#	91	0.00
46 T	Dibromomethane	0.188	0.174	7.4	92	0.00
47 T	Bromodichloromethane	0.478	0.443	7.3	90	0.00
48 T	Methyl methacrylate	0.221	0.219	0.9	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012524\
 Data File : VY017101.D
 Acq On : 25 Jan 2024 11:27
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICSVY012524

Quant Time: Jan 26 04:40:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012424S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 25 04:44:10 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	104	0.00
50 S	Toluene-d8	1.140	1.202	-5.4	94	0.00
51 T	4-Methyl-2-Pentanone	0.212	0.206	2.8	93	0.00
52 CM	Toluene	0.873	0.818	6.3#	90	0.00
53 T	t-1,3-Dichloropropene	0.437	0.425	2.7	93	0.00
54 T	cis-1,3-Dichloropropene	0.537	0.509	5.2	91	0.00
55 T	1,1,2-Trichloroethane	0.263	0.240	8.7	90	0.00
56 T	Ethyl methacrylate	0.230	0.228	0.9	92	0.00
57 T	1,3-Dichloropropane	0.444	0.416	6.3	90	0.00
58 T	2-Chloroethyl Vinyl ether	0.159	0.190	-19.5	101	0.00
59 T	2-Hexanone	0.162	0.145	10.5	93	0.00
60 T	Dibromochloromethane	0.305	0.281	7.9	91	0.00
61 T	1,2-Dibromoethane	0.239	0.226	5.4	93	0.00
62 S	4-Bromofluorobenzene	0.351	0.372	-6.0	94	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.349	0.308	11.7	89	0.00
65 PM	Chlorobenzene	1.060	0.956	9.8	90	0.00
66 T	1,1,1,2-Tetrachloroethane	0.387	0.354	8.5	89	0.00
67 C	Ethyl Benzene	1.926	1.791	7.0#	90	0.00
68 T	m/p-Xylenes	0.724	0.672	7.2	90	0.00
69 T	o-Xylene	0.657	0.626	4.7	91	0.00
70 T	Styrene	0.954	0.922	3.4	90	0.00
71 P	Bromoform	0.194	0.176	9.3	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
73 T	Isopropylbenzene	4.110	3.797	7.6	90	0.00
74 T	N-amyl acetate	0.829	0.791	4.6	91	0.00
75 P	1,1,2,2-Tetrachloroethane	0.784	0.721	8.0	91	0.00
76 T	1,2,3-Trichloropropane	0.607	0.551	9.2	86	0.00
77 T	Bromobenzene	0.889	0.818	8.0	90	0.00
78 T	n-propylbenzene	4.944	4.615	6.7	89	0.00
79 T	2-Chlorotoluene	2.713	2.508	7.6	90	0.00
80 T	1,3,5-Trimethylbenzene	3.253	3.038	6.6	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.234	0.219	6.4	94	0.00
82 T	4-Chlorotoluene	2.812	2.585	8.1	89	0.00
83 T	tert-Butylbenzene	2.814	2.559	9.1	86	0.00
84 T	1,2,4-Trimethylbenzene	3.233	3.007	7.0	89	0.00
85 T	sec-Butylbenzene	4.167	3.905	6.3	88	0.00
86 T	p-Isopropyltoluene	3.471	3.269	5.8	88	0.00
87 T	1,3-Dichlorobenzene	1.756	1.590	9.5	89	0.00
88 T	1,4-Dichlorobenzene	1.728	1.549	10.4	88	0.00
89 T	n-Butylbenzene	3.348	3.144	6.1	88	0.00
90 T	Hexachloroethane	0.723	0.651	10.0	88	0.00
91 T	1,2-Dichlorobenzene	1.500	1.375	8.3	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.100	0.092	8.0	92	0.00
93 T	1,2,4-Trichlorobenzene	0.784	0.699	10.8	87	0.00
94 T	Hexachlorobutadiene	0.472	0.402	14.8	85	0.00
95 T	Naphthalene	1.384	1.310	5.3	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012524\
Data File : VY017101.D
Acq On : 25 Jan 2024 11:27
Operator : SY/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY012524

Quant Time: Jan 26 04:40:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012424S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 25 04:44:10 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.652	0.575	11.8	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6