

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012622\
 Data File : VY007305.D
 Acq On : 26 Jan 2022 14:35
 Operator : SY/MD
 Sample : N1266-08
 Misc : 6.91g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RB-11-(2.10-3.6)

Quant Time: Jan 27 00:34:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	84995	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	158704	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	149755	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	64849	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	54391	52.154	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.300%
35) Dibromofluoromethane	7.722	113	52710	49.118	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.240%
50) Toluene-d8	10.179	98	193242	49.282	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.560%
62) 4-Bromofluorobenzene	12.483	95	71989	51.612	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012622\
Data File : VY007305.D
Acq On : 26 Jan 2022 14:35
Operator : SY/MD
Sample : N1266-08
Misc : 6.91g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RB-11-(2.10-3.6)

Quant Time: Jan 27 00:34:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 14 05:48:29 2022
Response via : Initial Calibration

