

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012622\
 Data File : VY007312.D
 Acq On : 26 Jan 2022 17:28
 Operator : SY/MD
 Sample : N1267-12
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ART-ROOM-FLOORS

Quant Time: Jan 27 00:35:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	80748	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	147161	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	130235	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	51895	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	44664	45.079	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	90.160%		
35) Dibromofluoromethane	7.722	113	46692	46.923	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	93.840%		
50) Toluene-d8	10.179	98	178460	49.082	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	98.160%		
62) 4-Bromofluorobenzene	12.483	95	55965	43.271	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	86.540%		
Target Compounds						
48) Methyl methacrylate	9.295	41	5290	7.328	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012622\
Data File : VY007312.D
Acq On : 26 Jan 2022 17:28
Operator : SY/MD
Sample : N1267-12
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ART-ROOM-FLOORS

Quant Time: Jan 27 00:35:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 14 05:48:29 2022
Response via : Initial Calibration

