

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012924\
 Data File : VY017127.D
 Acq On : 29 Jan 2024 11:27
 Operator : SY/MD
 Sample : VY0129SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0129SBL01

Quant Time: Jan 30 00:28:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012424S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 25 04:44:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	91022	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.710	114	192651	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.508	117	187428	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	71920	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	43733	49.314	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.620%
35) Dibromofluoromethane	7.734	113	46379	39.888	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	79.780%
50) Toluene-d8	10.197	98	171245	38.993	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	77.980%
62) 4-Bromofluorobenzene	12.495	95	54448	40.246	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	80.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012924\
Data File : VY017127.D
Acq On : 29 Jan 2024 11:27
Operator : SY/MD
Sample : VY0129SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0129SBL01

Quant Time: Jan 30 00:28:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012424S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 25 04:44:10 2024
Response via : Initial Calibration

