

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013124\
 Data File : VY017155.D
 Acq On : 31 Jan 2024 13:14
 Operator : SY/MD
 Sample : P1296-01
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-01-013024

Quant Time: Jan 31 22:58:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012424S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 25 04:44:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.813	168	81932	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.709	114	177798	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	175668	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	68634	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	39970	50.071	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.140%
35) Dibromofluoromethane	7.746	113	42067	39.202	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	78.400%
50) Toluene-d8	10.197	98	158345	39.068	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	78.140%
62) 4-Bromofluorobenzene	12.495	95	53681	42.994	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	85.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013124\
Data File : VY017155.D
Acq On : 31 Jan 2024 13:14
Operator : SY/MD
Sample : P1296-01
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-01-013024

Quant Time: Jan 31 22:58:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012424S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 25 04:44:10 2024
Response via : Initial Calibration

