

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013124\
 Data File : VY017162.D
 Acq On : 31 Jan 2024 17:13
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 31 23:00:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012424S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 25 04:44:10 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	72	0.00
2 T	Dichlorodifluoromethane	0.434	0.312	28.1#	65	0.00
3 P	Chloromethane	0.729	0.625	14.3	73	0.00
4 C	Vinyl Chloride	0.833	0.756	9.2#	75	0.00
5 T	Bromomethane	0.621	0.596	4.0	82	0.00
6 T	Chloroethane	0.534	0.540	-1.1	82	0.00
7 T	Trichlorofluoromethane	0.966	0.823	14.8	73	0.00
8 T	Diethyl Ether	0.288	0.294	-2.1	79	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.559	0.514	8.1	73	0.00
10 T	Methyl Iodide	0.531	0.467	12.1	62	0.00
11 T	Tert butyl alcohol	0.037	0.045	-21.6	102	0.00
12 CM	1,1-Dichloroethene	0.512	0.463	9.6#	71	0.00
13 T	Acrolein	0.058	0.052	10.3	67	0.00
14 T	Allyl chloride	0.774	0.764	1.3	75	0.00
15 T	Acrylonitrile	0.125	0.153	-22.4	92	0.00
16 T	Acetone	0.158	0.144	8.9	88	0.00
17 T	Carbon Disulfide	1.543	1.155	25.1#	58	0.00
18 T	Methyl Acetate	0.500	0.686	-37.2#	99	-0.01
19 T	Methyl tert-butyl Ether	1.249	1.387	-11.0	83	0.00
20 T	Methylene Chloride	0.671	0.609	9.2	80	0.00
21 T	trans-1,2-Dichloroethene	0.592	0.553	6.6	71	0.00
22 T	Diisopropyl ether	1.708	1.926	-12.8	83	0.00
23 T	Vinyl Acetate	0.829	0.913	-10.1	82	0.00
24 P	1,1-Dichloroethane	1.042	1.065	-2.2	79	0.00
25 T	2-Butanone	0.194	0.207	-6.7	92	0.00
26 T	2,2-Dichloropropane	0.890	0.868	2.5	75	0.00
27 T	cis-1,2-Dichloroethene	0.656	0.687	-4.7	79	0.00
28 T	Bromochloromethane	0.415	0.506	-21.9	79	0.00
29 T	Tetrahydrofuran	0.099	0.127	-28.3#	94	0.00
30 C	Chloroform	1.061	1.134	-6.9#	82	0.00
31 T	Cyclohexane	0.963	0.803	16.6	66	0.00
32 T	1,1,1-Trichloroethane	0.910	0.902	0.9	76	0.00
33 S	1,2-Dichloroethane-d4	0.487	0.538	-10.5	78	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	75	0.00
35 S	Dibromofluoromethane	0.302	0.311	-3.0	74	0.00
36 T	1,1-Dichloropropene	0.497	0.453	8.9	72	0.00
37 T	Ethyl Acetate	0.213	0.247	-16.0	94	0.00
38 T	Carbon Tetrachloride	0.487	0.449	7.8	74	0.00
39 T	Methylcyclohexane	0.597	0.518	13.2	66	0.00
40 TM	Benzene	1.442	1.416	1.8	76	0.00
41 T	Methacrylonitrile	0.114	0.150	-31.6#	113	0.00
42 TM	1,2-Dichloroethane	0.374	0.399	-6.7	84	0.00
43 T	Isopropyl Acetate	0.399	0.461	-15.5	89	0.00
44 TM	Trichloroethene	0.370	0.352	4.9	75	0.00
45 C	1,2-Dichloropropane	0.354	0.373	-5.4#	82	0.00
46 T	Dibromomethane	0.188	0.203	-8.0	85	0.00
47 T	Bromodichloromethane	0.478	0.509	-6.5	82	0.00
48 T	Methyl methacrylate	0.221	0.262	-18.6	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013124\
 Data File : VY017162.D
 Acq On : 31 Jan 2024 17:13
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 31 23:00:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012424S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 25 04:44:10 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	102	-0.01
50 S	Toluene-d8	1.140	1.172	-2.8	73	0.00
51 T	4-Methyl-2-Pentanone	0.212	0.269	-26.9#	97	0.00
52 CM	Toluene	0.873	0.885	-1.4#	78	0.00
53 T	t-1,3-Dichloropropene	0.437	0.474	-8.5	83	0.00
54 T	cis-1,3-Dichloropropene	0.537	0.563	-4.8	80	0.00
55 T	1,1,2-Trichloroethane	0.263	0.292	-11.0	87	0.00
56 T	Ethyl methacrylate	0.230	0.275	-19.6	88	0.00
57 T	1,3-Dichloropropane	0.444	0.493	-11.0	85	0.00
58 T	2-Chloroethyl Vinyl ether	0.159	0.210	-32.1#	89	0.00
59 T	2-Hexanone	0.162	0.188	-16.0	96	0.00
60 T	Dibromochloromethane	0.305	0.338	-10.8	86	0.00
61 T	1,2-Dibromoethane	0.239	0.265	-10.9	87	0.00
62 S	4-Bromofluorobenzene	0.351	0.377	-7.4	76	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	77	0.00
64 T	Tetrachloroethene	0.349	0.320	8.3	76	0.00
65 PM	Chlorobenzene	1.060	1.057	0.3	81	0.00
66 T	1,1,1,2-Tetrachloroethane	0.387	0.394	-1.8	82	0.00
67 C	Ethyl Benzene	1.926	1.909	0.9#	78	0.00
68 T	m/p-Xylenes	0.724	0.712	1.7	78	0.00
69 T	o-Xylene	0.657	0.662	-0.8	78	0.00
70 T	Styrene	0.954	1.029	-7.9	82	0.00
71 P	Bromoform	0.194	0.213	-9.8	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
73 T	Isopropylbenzene	4.110	3.857	6.2	79	0.00
74 T	N-amyl acetate	0.829	0.929	-12.1	93	0.00
75 P	1,1,2,2-Tetrachloroethane	0.784	0.840	-7.1	92	0.00
76 T	1,2,3-Trichloropropane	0.607	0.518	14.7	70	0.00
77 T	Bromobenzene	0.889	0.874	1.7	83	0.00
78 T	n-propylbenzene	4.944	4.742	4.1	79	0.00
79 T	2-Chlorotoluene	2.713	2.609	3.8	81	0.00
80 T	1,3,5-Trimethylbenzene	3.253	3.170	2.6	81	0.00
81 T	trans-1,4-Dichloro-2-butene	0.234	0.247	-5.6	92	0.00
82 T	4-Chlorotoluene	2.812	2.671	5.0	80	0.00
83 T	tert-Butylbenzene	2.814	2.693	4.3	79	0.00
84 T	1,2,4-Trimethylbenzene	3.233	3.131	3.2	80	0.00
85 T	sec-Butylbenzene	4.167	4.027	3.4	79	0.00
86 T	p-Isopropyltoluene	3.471	3.373	2.8	79	0.00
87 T	1,3-Dichlorobenzene	1.756	1.703	3.0	82	0.00
88 T	1,4-Dichlorobenzene	1.728	1.681	2.7	83	0.00
89 T	n-Butylbenzene	3.348	3.279	2.1	79	0.00
90 T	Hexachloroethane	0.723	0.687	5.0	81	0.00
91 T	1,2-Dichlorobenzene	1.500	1.500	0.0	86	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.100	0.112	-12.0	97	0.00
93 T	1,2,4-Trichlorobenzene	0.784	0.782	0.3	85	0.00
94 T	Hexachlorobutadiene	0.472	0.439	7.0	81	0.00
95 T	Naphthalene	1.384	1.569	-13.4	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013124\
Data File : VY017162.D
Acq On : 31 Jan 2024 17:13
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jan 31 23:00:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012424S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 25 04:44:10 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.652	0.673	-3.2	88	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6