

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013125\
 Data File : VY021003.D
 Acq On : 31 Jan 2025 11:46
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Jan 31 22:23:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y013025S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 31 16:50:06 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	235064	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	411416	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	366768	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	149862	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	122005	60.961	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.920%
35) Dibromofluoromethane	7.634	113	130277	52.672	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.340%
50) Toluene-d8	10.109	98	476887	49.399	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.800%
62) 4-Bromofluorobenzene	12.408	95	156246	48.234	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	96.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013125\
Data File : VY021003.D
Acq On : 31 Jan 2025 11:46
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Jan 31 22:23:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y013025S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 31 16:50:06 2025
Response via : Initial Calibration

