

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013125\
 Data File : VY021009.D
 Acq On : 31 Jan 2025 14:59
 Operator : SY/MD
 Sample : Q1239-04
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 348

Quant Time: Jan 31 22:25:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y013025S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 31 16:50:06 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	9943	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	13078	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	9108	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	1491	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	4667	55.129	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.260%
35) Dibromofluoromethane	7.640	113	4668	59.372	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	118.740%
50) Toluene-d8	10.115	98	13519	44.055	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	88.100%
62) 4-Bromofluorobenzene	12.414	95	2387	23.181	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	46.360%
Target Compounds						
16) Acetone	3.873	43	709	61.448	ug/l	Qvalue # 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013125\
Data File : VY021009.D
Acq On : 31 Jan 2025 14:59
Operator : SY/MD
Sample : Q1239-04
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
348

Quant Time: Jan 31 22:25:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y013025S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 31 16:50:06 2025
Response via : Initial Calibration

