

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020524\
 Data File : VY017225.D
 Acq On : 05 Feb 2024 13:43
 Operator : SY/MD
 Sample : P1367-01
 Misc : 5.23g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EB-6-0-2

Quant Time: Feb 06 00:58:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	84216	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.703	114	182249	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	181272	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	65472	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	41071	54.591	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.180%
35) Dibromofluoromethane	7.734	113	39291	39.929	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	79.860%
50) Toluene-d8	10.191	98	163449	45.621	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.240%
62) 4-Bromofluorobenzene	12.495	95	51864	44.190	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	88.380%
Target Compounds						
						Qvalue
16) Acetone	3.960	43	10947	49.507	ug/l	99
17) Carbon Disulfide	4.216	76	4845	1.859	ug/l	99
44) Trichloroethene	8.959	130	2303	1.872	ug/l	97
52) Toluene	10.258	92	6472	2.212	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020524\
Data File : VY017225.D
Acq On : 05 Feb 2024 13:43
Operator : SY/MD
Sample : P1367-01
Misc : 5.23g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EB-6-0-2

Quant Time: Feb 06 00:58:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 02 00:56:04 2024
Response via : Initial Calibration

