

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020524\
 Data File : VY017226.D
 Acq On : 05 Feb 2024 14:06
 Operator : SY/MD
 Sample : P1367-05
 Misc : 4.87g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EB-6-7-9

Quant Time: Feb 06 00:59:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	83150	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.704	114	179175	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	182335	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	72219	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	39628	53.349	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.700%
35) Dibromofluoromethane	7.734	113	42217	43.638	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	87.280%
50) Toluene-d8	10.191	98	157490	44.712	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.420%
62) 4-Bromofluorobenzene	12.489	95	53217	46.121	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	92.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020524\
Data File : VY017226.D
Acq On : 05 Feb 2024 14:06
Operator : SY/MD
Sample : P1367-05
Misc : 4.87g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EB-6-7-9

Quant Time: Feb 06 00:59:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 02 00:56:04 2024
Response via : Initial Calibration

