

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020524\
 Data File : VY017227.D
 Acq On : 05 Feb 2024 14:30
 Operator : SY/MD
 Sample : P1367-13
 Misc : 5.32g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EB-5-6-8

Quant Time: Feb 06 00:59:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	82703	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	176040	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.501	117	180598	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	71541	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.154	65	38934	52.698	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	105.400%
35) Dibromofluoromethane	7.734	113	42099	44.291	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.580%
50) Toluene-d8	10.191	98	155405	44.906	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.820%
62) 4-Bromofluorobenzene	12.489	95	53457	47.154	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	94.300%
Target Compounds						
					Qvalue	
16) Acetone	3.966	43	2859	13.166	ug/l	98
27) cis-1,2-Dichloroethene	6.996	96	1397	1.370	ug/l	98
44) Trichloroethene	8.953	130	5021	4.225	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020524\
Data File : VY017227.D
Acq On : 05 Feb 2024 14:30
Operator : SY/MD
Sample : P1367-13
Misc : 5.32g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EB-5-6-8

Quant Time: Feb 06 00:59:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 02 00:56:04 2024
Response via : Initial Calibration

