

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020524\
 Data File : VY017230.D
 Acq On : 05 Feb 2024 15:40
 Operator : SY/MD
 Sample : P1367-07
 Misc : 5.82g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EB-7-5-7

Quant Time: Feb 06 01:00:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	80202	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	173364	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	176339	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	77671	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	37847	52.824	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 105.640%	
35) Dibromofluoromethane	7.728	113	40576	43.348	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 86.700%	
50) Toluene-d8	10.191	98	154135	45.226	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 90.460%	
62) 4-Bromofluorobenzene	12.489	95	57242	51.272	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 102.540%	
Target Compounds						
						Qvalue
16) Acetone	3.991	43	5615	26.664	ug/l	88
73) Isopropylbenzene	12.343	105	6289	1.102	ug/l #	74
86) p-Isopropyltoluene	13.373	119	689931	142.009	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020524\
Data File : VY017230.D
Acq On : 05 Feb 2024 15:40
Operator : SY/MD
Sample : P1367-07
Misc : 5.82g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EB-7-5-7

Quant Time: Feb 06 01:00:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 02 00:56:04 2024
Response via : Initial Calibration

