

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020524\
 Data File : VY017231.D
 Acq On : 05 Feb 2024 16:04
 Operator : SY/MD
 Sample : P1307-11RE
 Misc : 4.23g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-614-VOC-02RE

Quant Time: Feb 06 01:01:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	83113	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	178700	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	184523	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	74211	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	42900	57.779	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.560%
35) Dibromofluoromethane	7.728	113	42632	44.184	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.360%
50) Toluene-d8	10.185	98	157191	44.746	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.500%
62) 4-Bromofluorobenzene	12.489	95	54682	47.517	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	95.040%
Target Compounds						
20) Methylene Chloride	4.728	84	12703	11.516	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020524\
Data File : VY017231.D
Acq On : 05 Feb 2024 16:04
Operator : SY/MD
Sample : P1307-11RE
Misc : 4.23g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-614-VOC-02RE

Quant Time: Feb 06 01:01:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 02 00:56:04 2024
Response via : Initial Calibration

