

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY020524\  
 Data File : VY017232.D  
 Acq On : 05 Feb 2024 16:27  
 Operator : SY/MD  
 Sample : P1362-07  
 Misc : 5.48g/5.0mL/MSVOA\_Y/SOIL/A  
 ALS Vial : 17 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 WC-2-VOC

Quant Time: Feb 06 01:01:40 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y020124S.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Feb 02 00:56:04 2024  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	7.795	168	4885	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	6594	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	3866	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	387	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	931	21.334	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	42.660%#
35) Dibromofluoromethane	7.728	113	1280	35.952	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	71.900%
50) Toluene-d8	10.185	98	5400	41.657	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	83.320%
62) 4-Bromofluorobenzene	12.489	95	545	12.834	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	25.660%#

Target Compounds	Qvalue
-----	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY020524\  
 Data File : VY017232.D  
 Acq On : 05 Feb 2024 16:27  
 Operator : SY/MD  
 Sample : P1362-07  
 Misc : 5.48g/5.0mL/MSVOA\_Y/SOIL/A  
 ALS Vial : 17 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 WC-2-VOC

Quant Time: Feb 06 01:01:40 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y020124S.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Feb 02 00:56:04 2024  
 Response via : Initial Calibration

