

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020524\
 Data File : VY017234.D
 Acq On : 05 Feb 2024 17:14
 Operator : SY/MD
 Sample : P1362-03
 Misc : 5.14g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-1-VOC

Quant Time: Feb 06 01:02:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	79187	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	170505	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	175035	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	67429	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	38124	53.893	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.780%
35) Dibromofluoromethane	7.722	113	41060	44.600	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.200%
50) Toluene-d8	10.185	98	152496	45.495	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.000%
62) 4-Bromofluorobenzene	12.483	95	50590	46.074	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	92.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020524\
Data File : VY017234.D
Acq On : 05 Feb 2024 17:14
Operator : SY/MD
Sample : P1362-03
Misc : 5.14g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-1-VOC

Quant Time: Feb 06 01:02:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 02 00:56:04 2024
Response via : Initial Calibration

