

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020524\
 Data File : VY017241.D
 Acq On : 05 Feb 2024 19:59
 Operator : SY/MD
 Sample : P1356-01
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-3-020224

Quant Time: Feb 06 01:05:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	64719	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	133505	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	133571	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	50411	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	28232	48.831	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.660%
35) Dibromofluoromethane	7.722	113	31916	44.276	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.560%
50) Toluene-d8	10.185	98	117133	44.630	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.260%
62) 4-Bromofluorobenzene	12.489	95	37226	43.299	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	86.600%
Target Compounds						
16) Acetone	3.954	43	2924	17.207	ug/l #	84
20) Methylene Chloride	4.722	84	6767	6.688	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020524\
Data File : VY017241.D
Acq On : 05 Feb 2024 19:59
Operator : SY/MD
Sample : P1356-01
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-3-020224

Quant Time: Feb 06 01:05:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 02 00:56:04 2024
Response via : Initial Calibration

