

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020524\
 Data File : VY017242.D
 Acq On : 05 Feb 2024 20:22
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 06 01:05:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	86	0.00
2 T	Dichlorodifluoromethane	0.379	0.344	9.2	91	0.00
3 P	Chloromethane	0.681	0.678	0.4	95	0.00
4 C	Vinyl Chloride	0.830	0.782	5.8#	91	0.00
5 T	Bromomethane	0.575	0.589	-2.4	100	0.00
6 T	Chloroethane	0.536	0.522	2.6	92	0.00
7 T	Trichlorofluoromethane	0.853	0.816	4.3	91	0.00
8 T	Diethyl Ether	0.268	0.273	-1.9	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.517	0.512	1.0	92	0.00
10 T	Methyl Iodide	0.544	0.478	12.1	78	0.00
11 T	Tert butyl alcohol	0.044	0.033	25.0	83	0.01
12 CM	1,1-Dichloroethene	0.485	0.485	0.0	93	0.00
13 T	Acrolein	0.041	0.036	12.2	89	0.00
14 T	Allyl chloride	0.733	0.752	-2.6	93	0.00
15 T	Acrylonitrile	0.121	0.120	0.8	90	0.00
16 T	Acetone	0.131	0.096	26.7#	73	0.00
17 T	Carbon Disulfide	1.547	1.587	-2.6	94	0.00
18 T	Methyl Acetate	0.293	0.331	-13.0	93	0.00
19 T	Methyl tert-butyl Ether	1.167	1.155	1.0	89	0.00
20 T	Methylene Chloride	0.617	0.587	4.9	97	0.00
21 T	trans-1,2-Dichloroethene	0.551	0.579	-5.1	95	0.00
22 T	Diisopropyl ether	1.609	1.689	-5.0	93	0.00
23 T	Vinyl Acetate	0.975	0.937	3.9	86	0.00
24 P	1,1-Dichloroethane	0.986	1.036	-5.1	96	0.00
25 T	2-Butanone	0.175	0.152	13.1	84	0.00
26 T	2,2-Dichloropropane	0.820	0.747	8.9	84	0.00
27 T	cis-1,2-Dichloroethene	0.617	0.646	-4.7	95	0.00
28 T	Bromochloromethane	0.377	0.381	-1.1	81	0.00
29 T	Tetrahydrofuran	0.099	0.099	0.0	91	0.00
30 C	Chloroform	0.992	1.023	-3.1#	94	0.00
31 T	Cyclohexane	0.900	0.880	2.2	92	0.00
32 T	1,1,1-Trichloroethane	0.853	0.867	-1.6	92	0.00
33 S	1,2-Dichloroethane-d4	0.447	0.453	-1.3	84	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	87	0.00
35 S	Dibromofluoromethane	0.270	0.275	-1.9	86	0.00
36 T	1,1-Dichloropropene	0.466	0.464	0.4	92	0.00
37 T	Ethyl Acetate	0.216	0.197	8.8	87	0.00
38 T	Carbon Tetrachloride	0.446	0.438	1.8	92	0.00
39 T	Methylcyclohexane	0.544	0.542	0.4	91	0.00
40 TM	Benzene	1.330	1.377	-3.5	97	0.00
41 T	Methacrylonitrile	0.102	0.095	6.9	89	0.00
42 TM	1,2-Dichloroethane	0.352	0.339	3.7	90	0.00
43 T	Isopropyl Acetate	0.382	0.368	3.7	90	0.00
44 TM	Trichloroethene	0.338	0.348	-3.0	95	0.00
45 C	1,2-Dichloropropane	0.327	0.335	-2.4#	95	0.00
46 T	Dibromomethane	0.176	0.174	1.1	91	0.00
47 T	Bromodichloromethane	0.442	0.445	-0.7	94	0.00
48 T	Methyl methacrylate	0.179	0.178	0.6	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020524\
 Data File : VY017242.D
 Acq On : 05 Feb 2024 20:22
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 06 01:05:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	87	0.00
50 S	Toluene-d8	0.983	1.043	-6.1	87	0.00
51 T	4-Methyl-2-Pentanone	0.205	0.203	1.0	91	0.00
52 CM	Toluene	0.803	0.833	-3.7#	94	0.00
53 T	t-1,3-Dichloropropene	0.415	0.404	2.7	89	0.00
54 T	cis-1,3-Dichloropropene	0.497	0.489	1.6	90	0.00
55 T	1,1,2-Trichloroethane	0.240	0.240	0.0	92	0.00
56 T	Ethyl methacrylate	0.292	0.302	-3.4	91	0.00
57 T	1,3-Dichloropropane	0.413	0.412	0.2	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.150	0.150	0.0	86	0.00
59 T	2-Hexanone	0.148	0.138	6.8	86	0.00
60 T	Dibromochloromethane	0.282	0.280	0.7	93	0.00
61 T	1,2-Dibromoethane	0.223	0.217	2.7	92	0.00
62 S	4-Bromofluorobenzene	0.322	0.328	-1.9	86	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	88	0.00
64 T	Tetrachloroethene	0.325	0.349	-7.4	103	0.00
65 PM	Chlorobenzene	0.965	0.961	0.4	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.351	0.350	0.3	94	0.00
67 C	Ethyl Benzene	1.743	1.800	-3.3#	95	0.00
68 T	m/p-Xylenes	0.662	0.680	-2.7	94	0.00
69 T	o-Xylene	0.609	0.627	-3.0	94	0.00
70 T	Styrene	1.004	1.060	-5.6	94	0.00
71 P	Bromoform	0.184	0.180	2.2	93	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
73 T	Isopropylbenzene	3.673	3.724	-1.4	95	0.00
74 T	N-amyl acetate	0.825	0.809	1.9	91	0.00
75 P	1,1,2,2-Tetrachloroethane	0.722	0.689	4.6	92	0.00
76 T	1,2,3-Trichloropropane	0.461	0.418	9.3	91	0.00
77 T	Bromobenzene	0.815	0.800	1.8	95	0.00
78 T	n-propylbenzene	4.511	4.566	-1.2	94	0.00
79 T	2-Chlorotoluene	2.480	2.477	0.1	95	0.00
80 T	1,3,5-Trimethylbenzene	2.921	2.984	-2.2	95	0.00
81 T	trans-1,4-Dichloro-2-butene	0.222	0.199	10.4	85	0.00
82 T	4-Chlorotoluene	2.553	2.493	2.4	92	0.00
83 T	tert-Butylbenzene	2.504	2.538	-1.4	94	0.00
84 T	1,2,4-Trimethylbenzene	2.884	2.936	-1.8	94	0.00
85 T	sec-Butylbenzene	3.914	3.953	-1.0	94	0.00
86 T	p-Isopropyltoluene	3.128	3.150	-0.7	94	0.00
87 T	1,3-Dichlorobenzene	1.625	1.615	0.6	96	0.00
88 T	1,4-Dichlorobenzene	1.579	1.556	1.5	97	0.00
89 T	n-Butylbenzene	2.974	2.983	-0.3	93	0.00
90 T	Hexachloroethane	0.642	0.644	-0.3	97	0.00
91 T	1,2-Dichlorobenzene	1.384	1.390	-0.4	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.094	0.088	6.4	93	0.00
93 T	1,2,4-Trichlorobenzene	0.687	0.727	-5.8	102	0.00
94 T	Hexachlorobutadiene	0.405	0.403	0.5	100	0.00
95 T	Naphthalene	1.270	1.385	-9.1	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020524\
Data File : VY017242.D
Acq On : 05 Feb 2024 20:22
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Feb 06 01:05:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 02 00:56:04 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.585	0.622	-6.3	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5