

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020722\
 Data File : VY007407.D
 Acq On : 07 Feb 2022 12:57
 Operator : SY/MD
 Sample : N1401-03
 Misc : 3.51g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PSC-124055

Quant Time: Feb 08 00:52:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 03 04:36:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	116087	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	210361	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	192232	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	74890	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	79194	59.309	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 118.620%	
35) Dibromofluoromethane	7.722	113	67003	49.602	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 99.200%	
50) Toluene-d8	10.185	98	256660	51.429	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 102.860%	
62) 4-Bromofluorobenzene	12.483	95	85563	50.443	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 100.880%	
Target Compounds						
16) Acetone	3.954	43	7059	16.278	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020722\
Data File : VY007407.D
Acq On : 07 Feb 2022 12:57
Operator : SY/MD
Sample : N1401-03
Misc : 3.51g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PSC-124055

Quant Time: Feb 08 00:52:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 03 04:36:48 2022
Response via : Initial Calibration

