

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020722\
 Data File : VY007408.D
 Acq On : 07 Feb 2022 13:20
 Operator : SY/MD
 Sample : N1401-05RE
 Misc : 3.75g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PSC-124042RE

Quant Time: Feb 08 00:37:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 03 04:36:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	57811	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	105014	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	96383	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	41195	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	38107	57.307	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 114.620%	
35) Dibromofluoromethane	7.722	113	33405	49.538	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 99.080%	
50) Toluene-d8	10.185	98	128754	51.681	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 103.360%	
62) 4-Bromofluorobenzene	12.483	95	43366	51.213	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 102.420%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020722\
Data File : VY007408.D
Acq On : 07 Feb 2022 13:20
Operator : SY/MD
Sample : N1401-05RE
Misc : 3.75g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PSC-124042RE

Quant Time: Feb 08 00:37:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 03 04:36:48 2022
Response via : Initial Calibration

