

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020722\
 Data File : VY007424.D
 Acq On : 07 Feb 2022 19:36
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 08 00:42:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 03 04:36:48 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	85	0.00
2 T	Dichlorodifluoromethane	0.585	0.465	20.5	70	0.00
3 P	Chloromethane	0.717	0.541	24.5	70	0.00
4 C	Vinyl Chloride	0.690	0.558	19.1#	69	0.00
5 T	Bromomethane	0.418	0.401	4.1	80	0.00
6 T	Chloroethane	0.424	0.370	12.7	74	0.00
7 T	Trichlorofluoromethane	1.119	0.941	15.9	71	0.00
8 T	Diethyl Ether	0.332	0.335	-0.9	86	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.630	0.539	14.4	73	0.00
10 T	Methyl Iodide	0.679	0.650	4.3	74	0.00
11 T	Tert butyl alcohol	0.092	0.067	27.2#	87	0.00
12 CM	1,1-Dichloroethene	0.582	0.497	14.6#	72	0.00
13 T	Acrolein	0.077	0.064	16.9	70	0.00
14 T	Allyl chloride	0.894	0.788	11.9	74	0.00
15 T	Acrylonitrile	0.156	0.171	-9.6	94	0.00
16 T	Acetone	0.187	0.158	15.5	62	0.00
17 T	Carbon Disulfide	1.827	1.525	16.5	70	0.00
18 T	Methyl Acetate	0.350	0.386	-10.3	96	0.00
19 T	Methyl tert-butyl Ether	1.640	1.719	-4.8	89	0.00
20 T	Methylene Chloride	0.653	0.601	8.0	82	0.00
21 T	trans-1,2-Dichloroethene	0.644	0.574	10.9	76	0.00
22 T	Diisopropyl ether	1.873	1.831	2.2	84	0.00
23 T	Vinyl Acetate	1.221	1.276	-4.5	88	0.00
24 P	1,1-Dichloroethane	1.131	1.040	8.0	78	0.00
25 T	2-Butanone	0.226	0.229	-1.3	80	0.00
26 T	2,2-Dichloropropane	1.082	0.906	16.3	71	0.00
27 T	cis-1,2-Dichloroethene	0.711	0.669	5.9	80	0.00
28 T	Bromochloromethane	0.442	0.448	-1.4	86	0.00
29 T	Tetrahydrofuran	0.130	0.150	-15.4	96	0.00
30 C	Chloroform	1.177	1.113	5.4#	81	0.00
31 T	Cyclohexane	1.122	0.886	21.0	70	0.00
32 T	1,1,1-Trichloroethane	1.094	0.971	11.2	75	0.00
33 S	1,2-Dichloroethane-d4	0.575	0.603	-4.9	89	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	87	0.00
35 S	Dibromofluoromethane	0.321	0.308	4.0	85	0.00
36 T	1,1-Dichloropropene	0.561	0.456	18.7	72	0.00
37 T	Ethyl Acetate	0.279	0.293	-5.0	94	0.00
38 T	Carbon Tetrachloride	0.590	0.491	16.8	72	0.00
39 T	Methylcyclohexane	0.705	0.573	18.7	71	0.00
40 TM	Benzene	1.488	1.315	11.6	78	0.00
41 T	Methacrylonitrile	0.139	0.141	-1.4	82	0.00
42 TM	1,2-Dichloroethane	0.452	0.437	3.3	85	0.00
43 T	Isopropyl Acetate	0.522	0.559	-7.1	93	0.00
44 TM	Trichloroethene	0.402	0.350	12.9	77	0.00
45 C	1,2-Dichloropropane	0.367	0.335	8.7#	81	0.00
46 T	Dibromomethane	0.218	0.215	1.4	87	0.00
47 T	Bromodichloromethane	0.528	0.503	4.7	83	0.00
48 T	Methyl methacrylate	0.236	0.250	-5.9	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020722\
 Data File : VY007424.D
 Acq On : 07 Feb 2022 19:36
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 08 00:42:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 03 04:36:48 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.003	-50.0#	97	0.00
50 S	Toluene-d8	1.186	1.105	6.8	82	0.00
51 T	4-Methyl-2-Pentanone	0.265	0.288	-8.7	94	0.00
52 CM	Toluene	0.937	0.828	11.6#	77	0.00
53 T	t-1,3-Dichloropropene	0.551	0.538	2.4	84	0.00
54 T	cis-1,3-Dichloropropene	0.618	0.588	4.9	82	0.00
55 T	1,1,2-Trichloroethane	0.287	0.284	1.0	87	0.00
56 T	Ethyl methacrylate	0.398	0.421	-5.8	91	0.00
57 T	1,3-Dichloropropane	0.508	0.502	1.2	87	0.00
58 T	2-Chloroethyl Vinyl ether	0.207	0.230	-11.1	96	0.00
59 T	2-Hexanone	0.193	0.200	-3.6	84	0.00
60 T	Dibromochloromethane	0.357	0.352	1.4	86	0.00
61 T	1,2-Dibromoethane	0.284	0.288	-1.4	88	0.00
62 S	4-Bromofluorobenzene	0.403	0.392	2.7	88	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	89	0.00
64 T	Tetrachloroethene	0.392	0.339	13.5	77	0.00
65 PM	Chlorobenzene	1.112	0.986	11.3	79	0.00
66 T	1,1,1,2-Tetrachloroethane	0.415	0.376	9.4	81	0.00
67 C	Ethyl Benzene	2.111	1.834	13.1#	78	0.00
68 T	m/p-Xylenes	0.782	0.677	13.4	76	0.00
69 T	o-Xylene	0.739	0.656	11.2	78	0.00
70 T	Styrene	1.218	1.135	6.8	82	0.00
71 P	Bromoform	0.229	0.230	-0.4	88	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	4.570	3.912	14.4	77	0.00
74 T	N-amyl acetate	1.131	1.208	-6.8	92	0.00
75 P	1,1,2,2-Tetrachloroethane	0.834	0.866	-3.8	93	0.00
76 T	1,2,3-Trichloropropane	0.594	0.643	-8.2	91	0.00
77 T	Bromobenzene	0.963	0.871	9.6	81	0.00
78 T	n-propylbenzene	5.535	4.758	14.0	76	0.00
79 T	2-Chlorotoluene	3.064	2.695	12.0	78	0.00
80 T	1,3,5-Trimethylbenzene	3.728	3.282	12.0	77	0.00
81 T	trans-1,4-Dichloro-2-butene	0.297	0.303	-2.0	87	0.00
82 T	4-Chlorotoluene	3.147	2.816	10.5	79	0.00
83 T	tert-Butylbenzene	3.245	2.817	13.2	77	0.00
84 T	1,2,4-Trimethylbenzene	3.670	3.259	11.2	78	0.00
85 T	sec-Butylbenzene	4.868	4.170	14.3	76	0.00
86 T	p-Isopropyltoluene	3.965	3.448	13.0	76	0.00
87 T	1,3-Dichlorobenzene	1.866	1.702	8.8	80	0.00
88 T	1,4-Dichlorobenzene	1.860	1.681	9.6	81	0.00
89 T	n-Butylbenzene	3.816	3.319	13.0	76	0.00
90 T	Hexachloroethane	0.749	0.667	10.9	78	0.00
91 T	1,2-Dichlorobenzene	1.643	1.524	7.2	83	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.136	0.147	-8.1	96	0.00
93 T	1,2,4-Trichlorobenzene	0.973	0.898	7.7	81	0.00
94 T	Hexachlorobutadiene	0.539	0.467	13.4	76	0.00
95 T	Naphthalene	2.022	2.112	-4.5	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020722\
Data File : VY007424.D
Acq On : 07 Feb 2022 19:36
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Feb 08 00:42:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 03 04:36:48 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.821	0.792	3.5	85	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6