

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020722\
 Data File : VY007402.D
 Acq On : 07 Feb 2022 09:54
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 02/08/2022
 Supervised By : Mahesh Dadoda 02/08/2022

Quant Time: Feb 08 00:34:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 03 04:36:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	140755	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	234130	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	205796	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	89534	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	74503	46.017	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	92.040%		
35) Dibromofluoromethane	7.728	113	65295	43.431	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	86.860%		
50) Toluene-d8	10.185	98	229019	41.232	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	82.460%		
62) 4-Bromofluorobenzene	12.483	95	81924	43.394	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	86.780%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.912	85	79754	48.396	ug/l	96
3) Chloromethane	2.119	50	89997	49.870	ug/l	96
4) Vinyl Chloride	2.259	62	94541	48.650	ug/l	99
5) Bromomethane	2.662	94	58661	49.836	ug/l	97
6) Chloroethane	2.802	64	57524	48.199	ug/l	98
7) Trichlorofluoromethane	3.137	101	153517	48.749	ug/l	92
8) Diethyl Ether	3.540	74	45570	48.751	ug/l	88
9) 1,1,2-Trichlorotrifluo...	3.912	101	85535	48.228	ug/l	93
10) Methyl Iodide	4.101	142	96049	50.236	ug/l	95
11) Tert butyl alcohol	4.954	59	38917	199.546	ug/l #	94
12) 1,1-Dichloroethene	3.887	96	79074	48.267	ug/l	88
13) Acrolein	3.735	56	45051	207.137	ug/l	97
14) Allyl chloride	4.491	41	127002	50.472	ug/l #	95
15) Acrylonitrile	5.174	53	111140	253.068	ug/l	99
16) Acetone	3.954	43	142756	271.493	ug/l	93
17) Carbon Disulfide	4.204	76	244429	47.525	ug/l	100
18) Methyl Acetate	4.485	43	49441	50.201	ug/l	92
19) Methyl tert-butyl Ether	5.228	73	230156	49.865	ug/l	97
20) Methylene Chloride	4.722	84	85566	50.220	ug/l	86
21) trans-1,2-Dichloroethene	5.235	96	86155	47.537	ug/l	92
22) Diisopropyl ether	6.125	45	261653	49.627	ug/l	96
23) Vinyl Acetate	6.070	43	866636	252.036	ug/l #	94
24) 1,1-Dichloroethane	6.027	63	155346	48.785	ug/l	98
25) 2-Butanone	6.990	43	168714	265.570	ug/l	91
26) 2,2-Dichloropropane	6.990	77	148691	48.800	ug/l	97
27) cis-1,2-Dichloroethene	6.996	96	94791	47.351	ug/l	93
28) Bromochloromethane	7.344	49	58995	47.382	ug/l	91
29) Tetrahydrofuran	7.356	42	94923	258.572	ug/l	90
30) Chloroform	7.515	83	160628	48.492	ug/l	98
31) Cyclohexane	7.795	56	146680	46.430	ug/l	94
32) 1,1,1-Trichloroethane	7.710	97	150153	48.772	ug/l	96
36) 1,1-Dichloropropene	7.923	75	127836	48.637	ug/l	99
37) Ethyl Acetate	7.082	43	66411	50.823	ug/l	97
38) Carbon Tetrachloride	7.905	117	134429	48.682	ug/l	99
39) Methylcyclohexane	9.191	83	159667	48.372	ug/l	93
40) Benzene	8.167	78	337841	48.492	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020722\
 Data File : VY007402.D
 Acq On : 07 Feb 2022 09:54
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 02/08/2022
 Supervised By : Mahesh Dadoda 02/08/2022

Quant Time: Feb 08 00:34:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 03 04:36:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.320	41	27766m	42.760	ug/l	
42) 1,2-Dichloroethane	8.246	62	105107	49.629	ug/l	96
43) Isopropyl Acetate	8.277	43	126387	51.719	ug/l #	93
44) Trichloroethene	8.947	130	91793	48.729	ug/l	94
45) 1,2-Dichloropropane	9.222	63	83782	48.770	ug/l	93
46) Dibromomethane	9.313	93	49014	47.955	ug/l	92
47) Bromodichloromethane	9.502	83	122950	49.746	ug/l	100
48) Methyl methacrylate	9.295	41	56544	51.164	ug/l	89
49) 1,4-Dioxane	9.307	88	11364	1015.035	ug/l	89
51) 4-Methyl-2-Pentanone	10.075	43	321376	259.306	ug/l	92
52) Toluene	10.252	92	211156	48.104	ug/l	96
53) t-1,3-Dichloropropene	10.471	75	130376	50.494	ug/l	100
54) cis-1,3-Dichloropropene	9.935	75	143373	49.575	ug/l #	87
55) 1,1,2-Trichloroethane	10.648	97	65984	49.156	ug/l	99
56) Ethyl methacrylate	10.514	69	96406	51.708	ug/l #	82
57) 1,3-Dichloropropane	10.795	76	116450	48.929	ug/l	97
58) 2-Chloroethyl Vinyl ether	9.789	63	244125	251.714	ug/l	97
59) 2-Hexanone	10.837	43	239574	265.147	ug/l	90
60) Dibromochloromethane	10.990	129	82883	49.597	ug/l	100
61) 1,2-Dibromoethane	11.093	107	66311	49.931	ug/l	99
64) Tetrachloroethene	10.727	164	77997	48.294	ug/l	95
65) Chlorobenzene	11.520	112	221509	48.385	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.593	131	82651	48.362	ug/l	97
67) Ethyl Benzene	11.599	91	418279	48.148	ug/l	100
68) m/p-Xylenes	11.709	106	308438	95.860	ug/l	99
69) o-Xylene	12.032	106	145644	47.876	ug/l	100
70) Styrene	12.050	104	244009	48.658	ug/l	99
71) Bromoform	12.215	173	47162	50.017	ug/l #	98
73) Isopropylbenzene	12.337	105	398608	48.713	ug/l	98
74) N-amyl acetate	12.148	43	105951	52.310	ug/l	92
75) 1,1,2,2-Tetrachloroethane	12.587	83	74097	49.586	ug/l	100
76) 1,2,3-Trichloropropane	12.636	75	53995m	50.728	ug/l	
77) Bromobenzene	12.611	156	82567	47.903	ug/l	95
78) n-propylbenzene	12.678	91	488737	49.312	ug/l	99
79) 2-Chlorotoluene	12.764	91	268951	49.012	ug/l	98
80) 1,3,5-Trimethylbenzene	12.819	105	326940	48.978	ug/l	97
81) trans-1,4-Dichloro-2-b...	12.380	75	27975	52.655	ug/l	87
82) 4-Chlorotoluene	12.861	91	273897	48.597	ug/l	98
83) tert-Butylbenzene	13.081	119	285214	49.080	ug/l	97
84) 1,2,4-Trimethylbenzene	13.123	105	319473	48.608	ug/l	99
85) sec-Butylbenzene	13.258	105	425817	48.853	ug/l	98
86) p-Isopropyltoluene	13.373	119	347048	48.877	ug/l	99
87) 1,3-Dichlorobenzene	13.373	146	161879	48.439	ug/l	98
88) 1,4-Dichlorobenzene	13.453	146	159277	47.812	ug/l	97
89) n-Butylbenzene	13.703	91	335397	49.077	ug/l	99
90) Hexachloroethane	13.965	117	65808	49.042	ug/l	79
91) 1,2-Dichlorobenzene	13.739	146	140021	47.601	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.361	75	11763	48.440	ug/l	90
93) 1,2,4-Trichlorobenzene	15.007	180	83460	47.910	ug/l	98
94) Hexachlorobutadiene	15.117	225	46315	48.028	ug/l	99
95) Naphthalene	15.239	128	181282	50.076	ug/l	99
96) 1,2,3-Trichlorobenzene	15.428	180	70481	47.931	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020722\
 Data File : VY007402.D
 Acq On : 07 Feb 2022 09:54
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/08/2022
 Supervised By :Mahesh Dadoda 02/08/2022

Quant Time: Feb 08 00:34:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 03 04:36:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020722\
Data File : VY007402.D
Acq On : 07 Feb 2022 09:54
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1


Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Quant Time: Feb 08 00:34:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 03 04:36:48 2022
Response via : Initial Calibration

Manual Integrations
APPROVED


Reviewed By : Semsettin Yesilyurt 02/08/2022
Supervised By : Mahesh Dadoda 02/08/2022





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.795 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:168 Resp: 14075
Ion Ratio Lower Upper
168 100
99 63.3 46.9 70.3




Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022






#2
Dichlorodifluoromethane
Concen: 48.396 ug/l
RT: 1.912 min Scan# 15
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54

Tgt Ion: 85 Resp: 79754
Ion Ratio Lower Upper
85 100
87 34.5 16.1 48.2



Abundance Scan 48 (2.113 min): VY007350.D\data.ms (-41)



#3

Chloromethane

Concen: 49.870 ug/l

RT: 2.119 min Scan# 49

Delta R.T. 0.006 min

Lab File: VY007402.D

Acq: 07 Feb 2022 09:54

Tgt Ion: 50 Resp: 8999
Ion Ratio Lower Upper
50 100
52 33.1 24.6 37.0

Instrument :

MSVOA_Y





ClientSampleId :

VSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022



#4

Vinyl Chloride

Concen: 48.650 ug/l


RT: 2.259 min Scan# 72


Delta R.T. 0.006 min

Lab File: VY007402.D

Acq: 07 Feb 2022 09:54


Tgt Ion: 62 Resp: 94541
Ion Ratio Lower Upper
62 100
64 31.7 25.0 37.6





#5
Bromomethane
Concen: 49.836 ug/l
RT: 2.662 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 94 Resp: 5866
Ion Ratio Lower Upper
94 100
96 95.0 73.5 110.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022



#6
Chloroethane
Concen: 48.199 ug/l
RT: 2.802 min Scan# 161
Delta R.T. 0.000 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54


Tgt Ion: 64 Resp: 57524
Ion Ratio Lower Upper
64 100
66 30.7 25.4 38.0





#7
 Trichlorofluoromethane
 Concen: 48.749 ug/l
 RT: 3.137 min Scan# 215
 Delta R.T. 0.006 min
 Lab File: VY007402.D
 Acq: 07 Feb 2022 09:54




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050




Tgt Ion:101 Resp: 15351
 Ion Ratio Lower Upper
 101 100
 103 65.8 48.0 72.0

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 02/08/2022
 Supervised By :Mahesh Dadoda 02/08/2022




#8
 Diethyl Ether
 Concen: 48.751 ug/l
 RT: 3.540 min Scan# 282
 Delta R.T. 0.006 min
 Lab File: VY007402.D
 Acq: 07 Feb 2022 09:54



Tgt Ion: 74 Resp: 45570
 Ion Ratio Lower Upper
 74 100
 45 98.1 55.1 165.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 48.228 ug/l
RT: 3.912 min Scan# 34
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 8553
Ion Ratio Lower Upper
101 100
85 45.8 40.6 60.8
151 74.7 64.9 97.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022



#10
Methyl Iodide
Concen: 50.236 ug/l
RT: 4.101 min Scan# 374
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54

Tgt Ion:142 Resp: 96049
Ion Ratio Lower Upper
142 100
127 47.9 41.5 62.3
141 15.0 11.9 17.9





#11

Tert butyl alcohol

Concen: 199.546 ug/l

RT: 4.954 min Scan# 513

Delta R.T. 0.006 min

Lab File: VY007402.D

Acq: 07 Feb 2022 09:54

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 59 Resp: 3891

Ion Ratio Lower Upper

59 100





57 10.0 6.4 9.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/08/2022

Supervised By :Mahesh Dadoda 02/08/2022



#12

1,1-Dichloroethene

Concen: 48.267 ug/l

RT: 3.887 min Scan# 339

Delta R.T. 0.006 min

Lab File: VY007402.D

Acq: 07 Feb 2022 09:54




Tgt Ion: 96 Resp: 79074


Ion Ratio Lower Upper

96 100

61 160.7 145.8 218.6

98 61.3 52.5 78.7





#13

Acrolein

Concen: 207.137 ug/l

RT: 3.735 min Scan# 314

Delta R.T. 0.000 min

Lab File: VY007402.D

Acq: 07 Feb 2022 09:54

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 56 Resp: 4505

Ion Ratio Lower Upper

56 100





55 69.1 57.6 86.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/08/2022

Supervised By :Mahesh Dadoda 02/08/2022



#14

Allyl chloride

Concen: 50.472 ug/l

RT: 4.491 min Scan# 438

Delta R.T. 0.006 min

Lab File: VY007402.D

Acq: 07 Feb 2022 09:54




Tgt Ion: 41 Resp: 127002


Ion Ratio Lower Upper

41 100

39 72.5 58.8 88.2


76 39.3 25.1 37.7#





#15
Acrylonitrile
Concen: 253.068 ug/l
RT: 5.174 min Scan# 511
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 53 Resp: 111140
Ion Ratio Lower Upper
53 100
52 83.3 67.0 100.4
51 36.8 29.8 44.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022




#16
Acetone
Concen: 271.493 ug/l
RT: 3.954 min Scan# 350
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Tgt Ion: 43 Resp: 142756
Ion Ratio Lower Upper
43 100
58 31.3 22.0 33.0





#17
Carbon Disulfide
Concen: 47.525 ug/l
RT: 4.204 min Scan# 391
Delta R.T. 0.000 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 76 Resp: 244429
Ion Ratio Lower Upper
76 100
78 9.2 7.4 11.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022




#18
Methyl Acetate
Concen: 50.201 ug/l
RT: 4.485 min Scan# 437
Delta R.T. 0.000 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Tgt Ion: 43 Resp: 49441
Ion Ratio Lower Upper
43 100
74 24.4 16.6 24.8





#19
Methyl tert-butyl Ether
Concen: 49.865 ug/l
RT: 5.228 min Scan# 51
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 73 Resp: 230156
Ion Ratio Lower Upper
73 100
57 21.9 18.6 27.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022




#20
Methylene Chloride
Concen: 50.220 ug/l
RT: 4.722 min Scan# 476
Delta R.T. 0.000 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Tgt Ion: 84 Resp: 85566
Ion Ratio Lower Upper
84 100
49 119.6 114.4 171.6
51 35.9 35.3 52.9
86 63.2 53.7 80.5





#21
trans-1,2-Dichloroethene
Concen: 47.537 ug/l
RT: 5.235 min Scan# 50
Delta R.T. 0.007 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 96 Resp: 8615
Ion Ratio Lower Upper
96 100
61 136.5 120.5 180.7
98 63.7 49.8 74.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022




#22
Diisopropyl ether
Concen: 49.627 ug/l
RT: 6.125 min Scan# 706
Delta R.T. 0.000 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Tgt Ion: 45 Resp: 261653
Ion Ratio Lower Upper
45 100
43 59.8 49.3 73.9
87 27.3 19.2 28.8
59 11.9 8.4 12.6





#23
 Vinyl Acetate
 Concen: 252.036 ug/l
 RT: 6.070 min Scan# 696
 Delta R.T. 0.006 min
 Lab File: VY007402.D
 Acq: 07 Feb 2022 09:54




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050




Tgt Ion: 43 Resp: 866630
 Ion Ratio Lower Upper
 43 100
 86 11.2 7.1 10.7

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 02/08/2022
 Supervised By :Mahesh Dadoda 02/08/2022




#24
 1,1-Dichloroethane
 Concen: 48.785 ug/l
 RT: 6.027 min Scan# 690
 Delta R.T. 0.006 min
 Lab File: VY007402.D
 Acq: 07 Feb 2022 09:54




Tgt Ion: 63 Resp: 155346
 Ion Ratio Lower Upper
 63 100
 98 6.5 3.6 10.8
 100 4.7 2.0 5.8





#25
2-Butanone
Concen: 265.570 ug/l
RT: 6.990 min Scan# 848
Delta R.T. 0.000 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 168714
Ion Ratio Lower Upper
43 100
72 26.2 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022



#26
2,2-Dichloropropane
Concen: 48.800 ug/l
RT: 6.990 min Scan# 848
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54


Tgt Ion: 77 Resp: 148691
Ion Ratio Lower Upper
77 100
97 22.5 10.4 31.4





#27
cis-1,2-Dichloroethene
Concen: 47.351 ug/l
RT: 6.996 min Scan# 848
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 96 Resp: 9479
Ion Ratio Lower Upper
96 100
61 148.3 0.0 319.8
98 65.7 0.0 127.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022



#28
Bromochloromethane
Concen: 47.382 ug/l
RT: 7.344 min Scan# 906
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54

Tgt Ion: 49 Resp: 58995
Ion Ratio Lower Upper
49 100
129 2.2 0.0 2.2
130 77.0 55.4 83.0





#29

Tetrahydrofuran

Concen: 258.572 ug/l

RT: 7.356 min Scan# 907

Delta R.T. 0.006 min

Lab File: VY007402.D

Acq: 07 Feb 2022 09:54

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion: 42 Resp: 9492

Ion	Ratio	Lower	Upper
42	100		
72	44.7	30.3	45.5
71	41.9	29.0	43.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022



#30

Chloroform

Concen: 48.492 ug/l

RT: 7.515 min Scan# 934




Delta R.T. 0.006 min


Lab File: VY007402.D

Acq: 07 Feb 2022 09:54

Tgt Ion: 83 Resp: 160628


Ion	Ratio	Lower	Upper
83	100		
85	62.8	51.8	77.6





#31
Cyclohexane
Concen: 46.430 ug/l
RT: 7.795 min Scan# 980
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 56 Resp: 146680
Ion Ratio Lower Upper
56 100
69 35.0 26.7 40.1
84 91.1 67.2 100.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022




#32
1,1,1-Trichloroethane
Concen: 48.772 ug/l
RT: 7.710 min Scan# 966
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54



Tgt Ion: 97 Resp: 150153
Ion Ratio Lower Upper
97 100
99 64.0 50.7 76.1
61 41.9 37.6 56.4





#33

1,2-Dichloroethane-d4

Concen: 46.017 ug/l

RT: 8.149 min Scan# 1037

Delta R.T. 0.006 min

Lab File: VY007402.D

Acq: 07 Feb 2022 09:54

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 65 Resp: 7450

Ion Ratio Lower Upper

65 100





67 53.1 0.0 99.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/08/2022

Supervised By :Mahesh Dadoda 02/08/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.697 min Scan# 1128

Delta R.T. 0.006 min

Lab File: VY007402.D

Acq: 07 Feb 2022 09:54




Tgt Ion:114 Resp: 234130


Ion Ratio Lower Upper

114 100

63 20.3 0.0 43.2


88 16.4 0.0 31.8





#35
Dibromofluoromethane
Concen: 43.431 ug/l
RT: 7.728 min Scan# 968
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 113 Resp: 6529
Ion Ratio Lower Upper
113 100
111 103.5 82.6 124.0
192 17.4 17.6 26.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022




#36
1,1-Dichloropropene
Concen: 48.637 ug/l
RT: 7.923 min Scan# 1001
Delta R.T. 0.000 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Tgt Ion: 75 Resp: 127836
Ion Ratio Lower Upper
75 100
110 33.7 17.1 51.2
77 30.9 24.4 36.6





#37
Ethyl Acetate
Concen: 50.823 ug/l
RT: 7.082 min Scan# 862
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 6641
Ion Ratio Lower Upper
43 100
61 13.5 9.7 14.5
70 9.6 7.1 10.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022



#38
Carbon Tetrachloride
Concen: 48.682 ug/l
RT: 7.905 min Scan# 998
Delta R.T. 0.000 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54

Tgt Ion:117 Resp: 134429
Ion Ratio Lower Upper
117 100
119 96.3 77.9 116.9
121 29.5 24.8 37.2





#39

Methylcyclohexane

Concen: 48.372 ug/l

RT: 9.191 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY007402.D

Acq: 07 Feb 2022 09:54

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 83 Resp: 15966

Ion Ratio Lower Upper

83 100

55 74.0 66.5 99.7





98 48.5 37.9 56.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/08/2022

Supervised By :Mahesh Dadoda 02/08/2022



#40

Benzene

Concen: 48.492 ug/l

RT: 8.167 min Scan# 1041

Delta R.T. 0.000 min

Lab File: VY007402.D




Acq: 07 Feb 2022 09:54


Tgt Ion: 78 Resp: 337841

Ion Ratio Lower Upper

78 100


77 24.3 20.4 30.6





#41
Methacrylonitrile
Concen: 42.760 ug/l m
RT: 7.320 min Scan# 901
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 41 Resp: 27760
Ion Ratio Lower Upper
41 100
39 0.0 92.9 139.3
67 0.0 0.0 0.0
52 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022



#42
1,2-Dichloroethane
Concen: 49.629 ug/l
RT: 8.246 min Scan# 1054
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54


Tgt Ion: 62 Resp: 105107
Ion Ratio Lower Upper
62 100
98 8.7 0.0 14.8





#43
Isopropyl Acetate
Concen: 51.719 ug/l
RT: 8.277 min Scan# 1059
Delta R.T. 0.000 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 43 Resp: 12638
Ion Ratio Lower Upper
43 100
61 28.4 20.2 30.2
87 13.6 8.9 13.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022




#44
Trichloroethene
Concen: 48.729 ug/l
RT: 8.947 min Scan# 1169
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54



Tgt Ion:130 Resp: 91793
Ion Ratio Lower Upper
130 100
95 99.6 0.0 188.2





#45

1,2-Dichloropropane

Concen: 48.770 ug/l

RT: 9.222 min Scan# 11

Delta R.T. 0.000 min

Lab File: VY007402.D

Acq: 07 Feb 2022 09:54

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 8378

Ion Ratio Lower Upper

63 100





65 32.4 23.0 34.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/08/2022

Supervised By :Mahesh Dadoda 02/08/2022



#46

Dibromomethane

Concen: 47.955 ug/l

RT: 9.313 min Scan# 1229

Delta R.T. 0.006 min

Lab File: VY007402.D

Acq: 07 Feb 2022 09:54




Tgt Ion: 93 Resp: 49014


Ion Ratio Lower Upper

93 100

95 83.7 64.3 96.5


174 93.1 83.4 125.0





#47
Bromodichloromethane
Concen: 49.746 ug/l
RT: 9.502 min Scan# 122950
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 83 Resp: 122950
Ion Ratio Lower Upper
83 100
85 64.6 51.9 77.9
127 8.4 6.6 10.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022




#48
Methyl methacrylate
Concen: 51.164 ug/l
RT: 9.295 min Scan# 1226
Delta R.T. 0.000 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54



Tgt Ion: 41 Resp: 56544
Ion Ratio Lower Upper
41 100
69 87.1 58.2 87.2
39 56.4 47.4 71.0





#49

1,4-Dioxane

Concen: 1015.035 ug/l

RT: 9.307 min Scan# 1136

Delta R.T. 0.006 min

Lab File: VY007402.D

Acq: 07 Feb 2022 09:54

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion: 88 Resp: 1136

Ion	Ratio	Lower	Upper
88	100		
43	34.6	32.8	49.2
58	65.0	59.8	89.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022



#50

Toluene-d8

Concen: 41.232 ug/l

RT: 10.185 min Scan# 1372




Delta R.T. 0.000 min


Lab File: VY007402.D

Acq: 07 Feb 2022 09:54

Tgt Ion: 98 Resp: 229019


Ion	Ratio	Lower	Upper
98	100		
100	65.8	50.2	75.2





#51
4-Methyl-2-Pentanone
Concen: 259.306 ug/l
RT: 10.075 min Scan# 1353
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 321370
Ion Ratio Lower Upper
43 100
58 40.0 28.1 42.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022



#52
Toluene
Concen: 48.104 ug/l
RT: 10.252 min Scan# 1383
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54


Tgt Ion: 92 Resp: 211156
Ion Ratio Lower Upper
92 100
91 171.9 142.6 213.8





#53
t-1,3-Dichloropropene
Concen: 50.494 ug/l
RT: 10.471 min Scan# 1418
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 75 Resp: 130370
Ion Ratio Lower Upper
75 100
77 31.7 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022



#54
cis-1,3-Dichloropropene
Concen: 49.575 ug/l
RT: 9.935 min Scan# 1331
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54


Tgt Ion: 75 Resp: 143373
Ion Ratio Lower Upper
75 100
77 30.9 25.6 38.4
39 45.3 47.6 71.4#





#55
1,1,2-Trichloroethane
Concen: 49.156 ug/l
RT: 10.648 min Scan# 1448
Delta R.T. 0.000 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 97 Resp: 65984
Ion Ratio Lower Upper
97 100
83 88.8 69.8 104.8
85 53.7 44.3 66.5
99 62.1 49.8 74.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022




#56
Ethyl methacrylate
Concen: 51.708 ug/l
RT: 10.514 min Scan# 1426
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54



Tgt Ion: 69 Resp: 96406
Ion Ratio Lower Upper
69 100
41 62.7 63.4 95.0#
39 34.9 35.8 53.8#









#57
1,3-Dichloropropane
Concen: 48.929 ug/l
RT: 10.795 min Scan# 1471
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 76 Resp: 116450
Ion Ratio Lower Upper
76 100
78 33.1 25.0 37.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022



#58
2-Chloroethyl Vinyl ether
Concen: 251.714 ug/l
RT: 9.789 min Scan# 1307
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54

Tgt Ion: 63 Resp: 244125
Ion Ratio Lower Upper
63 100
106 28.2 21.5 32.3









#59
2-Hexanone
Concen: 265.147 ug/l
RT: 10.837 min Scan# 1478
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 239574
Ion Ratio Lower Upper
43 100
58 55.7 24.4 73.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022



#60
Dibromochloromethane
Concen: 49.597 ug/l
RT: 10.990 min Scan# 1504
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54

Tgt Ion:129 Resp: 82883
Ion Ratio Lower Upper
129 100
127 77.0 38.7 116.1









#61
1,2-Dibromoethane
Concen: 49.931 ug/l
RT: 11.093 min Scan# 11093
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:107 Resp: 6631
Ion Ratio Lower Upper
107 100
109 93.4 75.4 113.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022



#62
4-Bromofluorobenzene
Concen: 43.394 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54


Tgt Ion: 95 Resp: 81924
Ion Ratio Lower Upper
95 100
174 74.8 0.0 177.0
176 72.0 0.0 175.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.496 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:117 Resp: 205790
Ion Ratio Lower Upper
117 100
82 56.3 43.0 64.4
119 31.5 25.4 38.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022




#64
Tetrachloroethene
Concen: 48.294 ug/l
RT: 10.727 min Scan# 1461
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Tgt Ion:164 Resp: 77997
Ion Ratio Lower Upper
164 100
166 129.1 100.4 150.6
129 95.8 71.4 107.0
131 93.2 70.8 106.2





#65
Chlorobenzene
Concen: 48.385 ug/l
RT: 11.520 min Scan# 1591
Delta R.T. 0.000 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:112 Resp: 221509
Ion Ratio Lower Upper
112 100
114 31.7 25.8 38.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022




#66
1,1,1,2-Tetrachloroethane
Concen: 48.362 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. 0.000 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54



Tgt Ion:131 Resp: 82651
Ion Ratio Lower Upper
131 100
133 94.1 47.9 143.6
119 64.7 30.8 92.3





#67

Ethyl Benzene

Concen: 48.148 ug/l

RT: 11.599 min Scan# 1604

Delta R.T. 0.006 min

Lab File: VY007402.D

Acq: 07 Feb 2022 09:54

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 91 Resp: 418279

Ion Ratio Lower Upper

91 100





106 30.7 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/08/2022

Supervised By :Mahesh Dadoda 02/08/2022



#68

m/p-Xylenes

Concen: 95.860 ug/l

RT: 11.709 min Scan# 1622

Delta R.T. 0.006 min

Lab File: VY007402.D




Acq: 07 Feb 2022 09:54


Tgt Ion:106 Resp: 308438

Ion Ratio Lower Upper

106 100

91 207.0 166.6 249.8









#69
o-Xylene
Concen: 47.876 ug/l
RT: 12.032 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:106 Resp: 145644
Ion Ratio Lower Upper
106 100
91 220.9 110.3 331.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022



#70
Styrene
Concen: 48.658 ug/l
RT: 12.050 min Scan# 1678
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54


Tgt Ion:104 Resp: 244009
Ion Ratio Lower Upper
104 100
78 51.2 42.4 63.6
103 55.1 44.4 66.6





#71
Bromoform
Concen: 50.017 ug/l
RT: 12.215 min Scan# 1705
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:173 Resp: 47162
Ion Ratio Lower Upper
173 100
175 48.1 24.6 74.0
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Tgt Ion:152 Resp: 89534
Ion Ratio Lower Upper
152 100
115 62.5 28.2 84.7
150 176.3 0.0 347.6





#73
Isopropylbenzene
Concen: 48.713 ug/l
RT: 12.337 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:105 Resp: 398608
Ion Ratio Lower Upper
105 100
120 25.7 13.4 40.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022




#74
N-ethyl acetate
Concen: 52.310 ug/l
RT: 12.148 min Scan# 1694
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54



Tgt Ion: 43 Resp: 105951
Ion Ratio Lower Upper
43 100
70 42.5 28.7 43.1
55 23.7 17.3 25.9
61 24.7 17.0 25.6









#75
1,1,2,2-Tetrachloroethane
Concen: 49.586 ug/l
RT: 12.587 min Scan# 1765
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 83 Resp: 7409
Ion Ratio Lower Upper
83 100
131 10.3 5.6 16.8
85 64.6 32.3 96.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022



#76
1,2,3-Trichloropropane
Concen: 50.728 ug/l m
RT: 12.636 min Scan# 1774
Delta R.T. 0.000 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54


Tgt Ion: 75 Resp: 53995
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77
Bromobenzene
Concen: 47.903 ug/l
RT: 12.611 min Scan# 1770
Delta R.T. 0.000 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:156 Resp: 8256
Ion Ratio Lower Upper
156 100
77 200.3 95.5 286.4
158 96.8 48.0 144.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022




#78
n-propylbenzene
Concen: 49.312 ug/l
RT: 12.678 min Scan# 1781
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54



Tgt Ion: 91 Resp: 488737
Ion Ratio Lower Upper
91 100
120 21.9 11.3 33.9









#79
2-Chlorotoluene
Concen: 49.012 ug/l
RT: 12.764 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 91 Resp: 26895
Ion Ratio Lower Upper
91 100
126 32.4 16.9 50.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022



#80
1,3,5-Trimethylbenzene
Concen: 48.978 ug/l
RT: 12.819 min Scan# 1804
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54


Tgt Ion:105 Resp: 326940
Ion Ratio Lower Upper
105 100
120 46.9 24.4 73.4





#81
trans-1,4-Dichloro-2-butene
Concen: 52.655 ug/l
RT: 12.380 min Scan# 1732
Delta R.T. 0.000 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 75 Resp: 27975
Ion Ratio Lower Upper
75 100
53 88.8 85.3 127.9
89 44.2 37.0 55.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022




#82
4-Chlorotoluene
Concen: 48.597 ug/l
RT: 12.861 min Scan# 1811
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Tgt Ion: 91 Resp: 273897
Ion Ratio Lower Upper
91 100
126 33.1 17.0 51.0





#83
tert-Butylbenzene
Concen: 49.080 ug/l
RT: 13.081 min Scan# 1846
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:119 Resp: 285214
Ion Ratio Lower Upper
119 100
91 66.2 34.8 104.4
134 25.7 13.3 39.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022



#84
1,2,4-Trimethylbenzene
Concen: 48.608 ug/l
RT: 13.123 min Scan# 1854
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54

Tgt Ion:105 Resp: 319473
Ion Ratio Lower Upper
105 100
120 43.9 22.3 66.9









#85
 sec-Butylbenzene
 Concen: 48.853 ug/l
 RT: 13.258 min Scan# 1876
 Delta R.T. 0.006 min
 Lab File: VY007402.D
 Acq: 07 Feb 2022 09:54

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Tgt Ion:105 Resp: 42581
 Ion Ratio Lower Upper
 105 100
 134 19.6 10.2 30.4




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 02/08/2022
 Supervised By :Mahesh Dadoda 02/08/2022



#86
 p-Isopropyltoluene
 Concen: 48.877 ug/l
 RT: 13.373 min Scan# 1895
 Delta R.T. 0.006 min
 Lab File: VY007402.D
 Acq: 07 Feb 2022 09:54

Tgt Ion:119 Resp: 347048
 Ion Ratio Lower Upper
 119 100
 134 25.3 13.2 39.5
 91 24.7 12.4 37.2





#87

1,3-Dichlorobenzene

Concen: 48.439 ug/l

RT: 13.373 min Scan# 1895

Delta R.T. 0.006 min

Lab File: VY007402.D

Acq: 07 Feb 2022 09:54

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:146 Resp: 161879

Ion Ratio Lower Upper

146 100

111 42.1 19.5 58.5





148 63.7 31.6 95.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/08/2022

Supervised By :Mahesh Dadoda 02/08/2022



#88

1,4-Dichlorobenzene

Concen: 47.812 ug/l

RT: 13.453 min Scan# 1908

Delta R.T. 0.006 min

Lab File: VY007402.D

Acq: 07 Feb 2022 09:54




Tgt Ion:146 Resp: 159277


Ion Ratio Lower Upper

146 100

111 41.4 19.3 57.8

148 63.5 32.6 97.8









#89
n-Butylbenzene
Concen: 49.077 ug/l
RT: 13.703 min Scan# 1949
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 91 Resp: 33539
Ion Ratio Lower Upper
91 100
92 52.9 26.3 78.9
134 24.7 13.0 38.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022



#90
Hexachloroethane
Concen: 49.042 ug/l
RT: 13.965 min Scan# 1992
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54


Tgt Ion:117 Resp: 65808
Ion Ratio Lower Upper
117 100
201 66.9 43.0 128.8





#91
1,2-Dichlorobenzene
Concen: 47.601 ug/l
RT: 13.739 min Scan# 1955
Delta R.T. 0.000 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:146 Resp: 14002
Ion Ratio Lower Upper
146 100
111 43.0 20.3 60.9
148 63.4 32.3 96.9

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022




#92
1,2-Dibromo-3-Chloropropane
Concen: 48.440 ug/l
RT: 14.361 min Scan# 2057
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54



Tgt Ion: 75 Resp: 11763
Ion Ratio Lower Upper
75 100
155 81.6 38.0 113.9
157 107.9 47.8 143.3





#93

1,2,4-Trichlorobenzene

Concen: 47.910 ug/l

RT: 15.007 min Scan# 2162

Delta R.T. 0.006 min

Lab File: VY007402.D

Acq: 07 Feb 2022 09:54

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:180 Resp: 83460

Ion Ratio Lower Upper

180 100

182 94.9 47.3 141.9





145 32.2 14.6 43.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/08/2022

Supervised By :Mahesh Dadoda 02/08/2022



#94

Hexachlorobutadiene

Concen: 48.028 ug/l

RT: 15.117 min Scan# 2181

Delta R.T. 0.006 min

Lab File: VY007402.D

Acq: 07 Feb 2022 09:54




Tgt Ion:225 Resp: 46315


Ion Ratio Lower Upper

225 100

223 62.3 30.8 92.3


227 64.2 31.6 94.7





#95
Naphthalene
Concen: 50.076 ug/l
RT: 15.239 min Scan# 21
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:128 Resp: 18128
Ion Ratio Lower Upper
128 100
127 13.1 11.0 16.4
129 10.7 8.4 12.6

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/08/2022
Supervised By :Mahesh Dadoda 02/08/2022



#96
1,2,3-Trichlorobenzene
Concen: 47.931 ug/l
RT: 15.428 min Scan# 2232
Delta R.T. 0.006 min
Lab File: VY007402.D
Acq: 07 Feb 2022 09:54



Tgt Ion:180 Resp: 70481
Ion Ratio Lower Upper
180 100
182 94.0 47.3 141.9
145 34.7 15.3 45.9

