

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020822\
 Data File : VY007432.D
 Acq On : 08 Feb 2022 14:46
 Operator : SY/MD
 Sample : N1432-07
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SP-3

Quant Time: Feb 09 00:55:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 03 04:36:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	102980	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	193197	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	187862	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	79854	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	68074	57.470	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.940%
35) Dibromofluoromethane	7.722	113	63798	51.426	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.860%
50) Toluene-d8	10.179	98	234694	51.205	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.420%
62) 4-Bromofluorobenzene	12.477	95	85571	54.929	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	109.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020822\
Data File : VY007432.D
Acq On : 08 Feb 2022 14:46
Operator : SY/MD
Sample : N1432-07
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SP-3

Quant Time: Feb 09 00:55:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 03 04:36:48 2022
Response via : Initial Calibration

