

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020822\
 Data File : VY007437.D
 Acq On : 08 Feb 2022 16:43
 Operator : SY/MD
 Sample : N1415-01RE
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CL-02-020422RE

Quant Time: Feb 09 00:58:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 03 04:36:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	41423	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	71709	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	61400	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	22227	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	21109	44.303	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.600%
35) Dibromofluoromethane	7.722	113	22868	49.662	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.320%
50) Toluene-d8	10.179	98	86153	50.642	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.280%
62) 4-Bromofluorobenzene	12.477	95	25096	43.402	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	86.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020822\
Data File : VY007437.D
Acq On : 08 Feb 2022 16:43
Operator : SY/MD
Sample : N1415-01RE
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CL-02-020422RE

Quant Time: Feb 09 00:58:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 03 04:36:48 2022
Response via : Initial Calibration

