

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020922\
 Data File : VY007459.D
 Acq On : 09 Feb 2022 17:43
 Operator : SY/MD
 Sample : N1444-01
 Misc : 4.45g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ERMIR-TPS

Quant Time: Feb 10 05:37:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 03 04:36:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	111926	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	209538	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	183192	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	61215	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	73079	56.764	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 113.520%	
35) Dibromofluoromethane	7.722	113	68415	50.847	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.700%	
50) Toluene-d8	10.185	98	250972	50.487	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 100.980%	
62) 4-Bromofluorobenzene	12.483	95	80855	47.854	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 95.700%	
Target Compounds						
16) Acetone	3.960	43	4505	10.774	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020922\
Data File : VY007459.D
Acq On : 09 Feb 2022 17:43
Operator : SY/MD
Sample : N1444-01
Misc : 4.45g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ERMR-TPS

Quant Time: Feb 10 05:37:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 03 04:36:48 2022
Response via : Initial Calibration

