



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020922\
Data File : VY007440.D
Acq On : 09 Feb 2022 08:52
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
Title : SW846 8260
Last Update : Thu Feb 03 04:36:48 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1741

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.9	9385	PASS
75	95	30	60	54.2	25568	PASS
95	95	100	100	100.0	47155	PASS
96	95	5	9	6.9	3268	PASS
173	174	0.00	2	1.1	397	PASS
174	95	50	100	76.7	36181	PASS
175	174	5	9	7.2	2621	PASS
176	174	95	101	95.5	34544	PASS
177	176	5	9	6.8	2343	PASS