

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020924\
 Data File : VY017337.D
 Acq On : 09 Feb 2024 14:20
 Operator : SY/MD
 Sample : P1414-06
 Misc : 6.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-9(10-11)

Quant Time: Feb 09 22:03:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	83954	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	162039	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	147872	50.000	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	13.428	152	57686	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	43207	57.610	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.220%
35) Dibromofluoromethane	7.716	113	41566	47.509	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.020%
50) Toluene-d8	10.179	98	141496	44.419	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	88.840%
62) 4-Bromofluorobenzene	12.483	95	41072	39.360	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	78.720%

Target Compounds						Qvalue
16) Acetone	3.954	43	3603	16.345	ug/l	# 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020924\
Data File : VY017337.D
Acq On : 09 Feb 2024 14:20
Operator : SY/MD
Sample : P1414-06
Misc : 6.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-9(10-11)

Quant Time: Feb 09 22:03:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 02 00:56:04 2024
Response via : Initial Calibration

