

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020924\
 Data File : VY017338.D
 Acq On : 09 Feb 2024 14:43
 Operator : SY/MD
 Sample : P1414-05
 Misc : 4.49g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-8(10-11)

Quant Time: Feb 09 22:03:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	84049	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	163409	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.490	117	147111	50.000	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	13.428	152	57933	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	41458	55.215	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.440%
35) Dibromofluoromethane	7.716	113	40754	46.190	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.380%
50) Toluene-d8	10.179	98	142404	44.330	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	88.660%
62) 4-Bromofluorobenzene	12.483	95	40727	38.702	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	77.400%
Target Compounds						Qvalue
16) Acetone	3.942	43	3164	14.337	ug/l	93
20) Methylene Chloride	4.710	84	8407	6.235	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020924\
Data File : VY017338.D
Acq On : 09 Feb 2024 14:43
Operator : SY/MD
Sample : P1414-05
Misc : 4.49g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-8(10-11)

Quant Time: Feb 09 22:03:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 02 00:56:04 2024
Response via : Initial Calibration

